

Review

Avogadro and Planck Constants, Two Pillars of the International System of Units

Enrico Massa 

INRiM—Istituto Nazionale di Ricerca Metrologica, Strada delle Cacce 91, 10135 Torino, Italy; e.massa@inrim.it

Abstract: The International System of Units (SI), the current form of the metric system and the world's most used system of units, has been continuously updated and refined since the Metre Convention of 1875 to ensure that it remains up to date with the latest scientific and technological advances. The General Conference on Weights and Measures, at its 26th meeting in 2018, decided to adopt stipulated values of seven physical constants linked to seven measurement units (the second, meter, kilogram, ampere, kelvin, mole, and candela). This paper reviews the technologies developed, in intense and long-standing work, to determine the Avogadro and Planck constants, which are now integral to realising the kilogram.

Keywords: International System of Units (SI); fundamental constants; kilogram; Kibble balance; XRCd method

1. Introduction

In ancient times, various civilizations developed methods for measuring physical quantities, such as mass, length, volume, and time. The ancient Egyptians developed beam balances, which were depicted in the *Book of the Dead* [1]. This book describes the “weighing of the heart” or “judgment of Osiris”. According to Egyptian mythology, after a person dies, the fate in the afterlife is determined by weighing the heart against a feather, where the feather acts as the reference standard for the unit of measure. If the heart is heavier than the feather, Ammit (the swallower of the dead) will eat the heart, and the death will be forever. If the scale remains balanced, the soul will join Osiris, the king of the blessed dead, for eternity.

Outside of mythology, for unit systems to be usable in everyday life, trade, and science, the units need to be standardized. This process began by using locally available materials and artefacts, such as body parts like feet and hands, as length standards. The metric system, which was established during the French Revolution, used natural quantities to create a system of units that would apply to all people, for all time. The meter was defined as the length of a meridian arc, and the kilogram was defined as the mass of one cubic decimeter of pure water at its maximum density. To ensure stability and practical realization, platinum prototypes of the meter and kilogram were created in 1799 and stored as references in the Archives de la République in Paris.

To implement a worldwide common system of units, on the 20 May 1875, the representatives of seventeen nations signed an international treaty commonly known as the Metre Convention [2]. The Metre Convention established the International Bureau of Weights and Measures (commonly referred to as BIPM by its French name), in Sèvres, France, charged with maintaining new international prototypes, the meter, (which—as with the Metre of the Archives—was significantly easy to use and durable) and the kilogram and to act as the repository of the physical measurement standards. The new kilogram, whose mass was essentially the same as the Kilogram of the Archives, was made from an alloy of platinum and iridium from a batch of material provided by Johnson, Matthey & Co., London, UK in 1880 [3] and called the International Prototype of the Kilogram [4] (IPK, also referred to by the metrologists as Le Grand K).



Citation: Massa, E. Avogadro and Planck Constants, Two Pillars of the International System of Units. *Physics* **2024**, *6*, 845–858. <https://doi.org/10.3390/physics6020052>

Received: 29 February 2024

Revised: 23 April 2024

Accepted: 25 April 2024

Published: 3 June 2024



Copyright: © 2024 by the author. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<https://creativecommons.org/licenses/by/4.0/>).

Before 1889, there was no official definition for the kilogram. It was only in that year, during the 1st General Conference on Weights and Measures (CGPM, Conférence Générale des Poids et Mesures, in French), that the mass of the IPK was established as being equal to 1 kg. The CGPM sanctioned that “this prototype shall henceforth be considered to be the unit of mass” [5]. Since then, this prototype has defined the kilogram and is kept in a BIPM’s vault. In 1901, the 3rd CGPM confirmed that “the kilogram is the unit of mass; it is equal to the mass of the International Prototype of the Kilogram” [6]. This statement was intended to end the ambiguity of using the word weight for mass. Because of these definitions, the mass of the international prototype is exactly equal to 1 kg and raised to the status of a universal constant.

After three verifications were carried out in 1889, 1946, and 1991 [7], the mass difference between the international prototype and its official copies, known as “*témions*”, has been observed to drift by about 50 μg per century. However, instead of all *témions* and many national copies drifting similarly, the observations pointed to the instability of the mass of the international prototype itself, despite being defined as invariant. The kilogram definition by an artefact caused concern among metrologists, who seek a new definition that allows for the measurement of the prototype mass’s changes.

The masses of *témions* were remeasured against the international prototype in 2014 [8] in anticipation of the redefinition of the kilogram (see Figure 1). This has resulted in the discovery of a 35 μg offset between the BIPM “as-maintained” mass unit and the mass of the international prototype. This offset further highlights the inadequacy of a definition based on a material artefact.




Figure 1. Mass variation in the official copies 32, 8₄₁, 7, 43, K1, and 47 to the mass of the International Prototype of the Kilogram. The official copy number, 8, was erroneously marked as 41. The measurements were carried out in 1889, 1946, 1991, and 2014 [8].

Although a variety of processes (e.g., surface contamination, alloying, and the diffusion of included hydrogen) might explain the mass drift, its reason is still experimentally unexplained. To tackle this issue, metrologists have come up with two (redundant and independent) methods of realizing the mass unit: the Kibble balance (KB, previously known as watt balance), linking the kilogram to the Planck constant; and the X-Ray Crystal Density (XRCD) method, linking the kilogram to the Avogadro constant. These two measurement methods have, since the 1980s, played a crucial role in the effort to redefine

the International System of Units (SI, *Système International d'unités*, in French), aligning it with our understanding of how nature behaves.

2. From Mass Artefacts to Physical Constants

In 1870, following the footsteps of the French revolutionaries, James Clerk Maxwell [9] suggested that physical units should not be founded on macroscopic quantities but on the properties of these “imperishable and unalterable” molecules. Later in 1899, Max Planck proposed a system of units that relied solely on universal constants [10].

The vision of the French revolutionaries, Maxwell and Planck to have a system of units based on natural constants was finally realized on 20 May 2019. Quantum mechanics and the semiconductor industry are deeply involved in replacing the kilogram artefact with realizations based on fundamental constants of physics.

In 1963, Claudio Egidi [11] proposed the idea of a mass standard that could be established using a crystal made up of a known number of atoms. He envisioned a perfect cubic crystal where the atoms are arranged regularly in the crystal lattice. By calculating the ratio of the macroscopic volume to the volume of a single atom, it is possible to determine the number of atoms. Once the atom had been established as the mass standard, the mass of the cube could be determined.

In 1965, Ulrich Bonse and Michael Hart [12] operated the first X-ray interferometer and paved the way for the achievement of Egidi's dream. Soon, Richard Deslattes and colleagues [13] used an X-ray interferometer to measure the lattice parameter of a silicon crystal, completed the first count of the atoms in a natural silicon crystal, and established a link between the kilogram and the Avogadro constant. To bypass limitations of the measurement accuracy due to the indirect determination of the silicon density and natural silicon [14] occurring in three isotopes, Gianfranco Zosi proposed to count the atoms in a sphere highly enriched with the isotope 28 and demonstrated that polishing perfect silicon spheres is possible [15,16].

To ensure that no discontinuity is produced to the 1901 definition, it was necessary for the mass of the ^{28}Si atom and, equivalently, for the Avogadro constant, to be more accurately measured in terms of the mass of the Pt–Ir prototype. Next, by reversing the experiment, the mass can be determined in terms of a (suitably) fixed value of the Avogadro constant.

In 2004, following Egidi and Zosi, a group of metrology laboratories, including the BIPM, INRiM (Italy), IRMM (Belgium), NMIA (Australia), NMIJ (Japan), NPL (UK), and PTB (Germany), combined their resources and expertise by launching an International Avogadro Coordination (IAC) to grow a silicon crystal enriched with ^{28}Si . The goal was twofold: firstly, to accurately determine the Avogadro constant, N_A [17–21]; secondly, to demonstrate that realizing a mass standard by counting ^{28}Si atoms was technologically possible.

In addition to the enrichment and growth of a hyper-pure and perfect crystal, several critical technologies needed to be developed *ex novo*. X-ray and optical interferometry were pushed to their limits, first by Giovanni Mana and colleagues [22] and then by the author of the present paper and colleagues [23]. The PTB and NMIJ laboratories developed new ways to measure the volume of silicon balls and characterize their surface layers in terms of thickness and chemical compositions [24,25]. Eventually, to determine the isotopic composition of the enriched crystal, PTB laboratory invented a new procedure of trace element analysis [26]. The IAC results are reported in a special issue of *Metrologia* published in 2011 [27].

To link a physical constant, precisely the Planck one, to the unit of mass, there is a second method. In 1976, Bryan Kibble developed a new way to realize the SI ampere [28]. This apparatus allowed electrical metrology to be accurately based on conventional electrical units but set it apart from the SI. By operating the “ampere current balance” in the reverse mode, the Kibble apparatus was subsequently perfected and evolved by Ian Robinson at the NPL [29], by Stephan Schlamming and colleagues at the NIST [30], by Carlos Sanchez and colleagues at the NRC [31], and by Matthieu Thomas and colleagues at the LNE [32] into a balance comparing mechanical and electrical powers. As in the case of the XRCD

experiment, before a kilogram realization based on this balance was possible, it was used to measure the Planck constant, h , in terms of the 1901 definition of the kilogram. Nowadays, it is run in reverse, realizes the mass unit from the Planck constant, and was renamed, firstly, watt balance and, next, Kibble balance. It is worth noting that the Rydberg, fine structure, Planck, and Avogadro constants— R_∞ , α , h , and N_A , respectively,—the speed of light, c , and the molar mass, M_e , of the electron, are interrelated by [33]

$$2R_\infty hc = \frac{\alpha^2 M_e c^2}{N_A}. \quad (1)$$

This relationship implies that the accurate determinations of R_∞ , α , and M_e determine the molar Planck constant as

$$N_A h = \frac{\alpha^2 M_e c}{2R_\infty}. \quad (2)$$

The result of the XRCD experiment delivered, de facto, also an accurate value of the Planck constant. Vice versa, the Kibble balance delivered, de facto, is also an accurate value of the Avogadro constant.

The “tour de force” for the determination of Avogadro and Planck constants sped up in 2005, after the publication of a paper [34] claiming that the “time for the redefinition of the kilogram had come”.

The CGPM was already looking at the redefinition of the kilogram. In 1999, the 21st CGPM recommended that metrologists “continue their efforts to refine experiments that link the unit of mass to fundamental or atomic constants. . .” (Resolution 7: The definition of the kilogram [35]). In 2007, the 23rd CGPM recommended that metrologists make it possible “to redefine the kg, the ampere, the kelvin, and the mole using fixed values of the fundamental constants at the time of the next 24th CGPM” (Resolution 12: On the possible redefinition of certain base units of the International System of Units (SI) [36]). At that time, the main impediment to redefining the kilogram was the discrepancy of about one part in 10^6 between the results of the Planck and Avogadro constants measurements, which were compared via the well-measured molar Planck constant value.

After significant progress in Kibble balance technology and XRCD experiments, in 2011, the 24th CGPM established the bases of the revised SI and proposed wording for the definitions of the base units (Resolution 1: On the possible future revision of the International System of Units, the SI [37]). In 2014, the 25th CGPM asked “to complete all work necessary for the CGPM at its 26th Meeting to adopt a resolution that would replace the current SI with the revised SI, provided the amount of data, their uncertainties, and level of consistency are deemed satisfactory” (Resolution 1: On the future revision of the International System of Units, the SI [38]).

At the same time, the Consultative Committee for Mass and Related Quantities (CCM) defined the requirements and, jointly with the Consultative Committee for Units (CCU) of the International Committee for Weights and Measures (CIPM, Comité International des Poids et Mesures, in French), the roadmap to make the redefinition of the kilogram possible [39].

The metrologists and CCM aimed for both agreement and low uncertainties of the measured values of the Planck and Avogadro constants. To achieve this, the CCM made a recommendation “On a new definition of the kilogram” in its 14th Meeting [40]. This recommendation put forward four conditions.

The first condition is that at least three independent experiments, including the Kibble balance and XRCD one, should yield consistent results of the measurements of the Planck and Avogadro constants with relative standard uncertainties not larger than 5 parts in 10^8 .

The second condition states that the uncertainty associated with at least one measurement result should not be larger than 2 parts in 10^8 .

The third condition requires that the measurement results be traceable to the masses of the BIPM’s mass standards and the international prototype of the kilogram.

The fourth condition is to validate the “mise en pratique” [41] of the new kilogram definition following the principles of the CIPM-MRA (Mutual Recognition Arrangement).

These requirements were successfully fulfilled at the time of the 16th Meeting of the CCM in 2017. As a result, the scientific work aimed at redefining the kilogram was concluded and the CGPM, at its 26th Meeting, rebuilt the SI. The redefinition led to the kilogram, ampere, kelvin, and mole, which were previously defined in terms of material artefacts or properties, being redefined through their linkage to stipulated values of fundamental constants of physics.

3. The Realization of the Kilogram

In the new SI, the kilogram is defined in terms of the Planck constant, which was previously determined (in terms of the 1901 definition of the kilogram) via the Kibble balance and XRCD experiment. The formal kilogram definition now reads [41]: “The kilogram, with the symbol kg, is the SI unit of mass. It is defined by taking the fixed numerical value of the Planck constant h to be $h = 6.62607015 \times 10^{-34}$ when expressed in the unit J s, which is equal to $\text{m}^2 \text{kg s}^{-1}$, where the meter and the second are defined in terms of c and $\Delta\nu_{\text{Cs}}$ ”.

Therefore, the Planck constant is conventionally stipulated to be $h = 6.62607015 \times 10^{-34}$ J s, where c and $\Delta\nu_{\text{Cs}}$ are the speed of light and the unperturbed ground-state hyperfine transition frequency of the ^{133}Cs atom, respectively. The numerical value of the Planck constant sets the size of the kilogram. It was accurately chosen so that no gap could be detected between the new and past realizations, within the associated uncertainties. This value was the result of a least-squares adjustment (LSA) of the values of the fundamental physical constants provided in 2017 by the Committee on Data for Science and Technology (CODATA) through its Task Group on Fundamental Physical Constants. Figure 2 summarizes the best determination of the Planck constant. It shows the results of the measurements carried out from 2011 to 2017 by both the Kibble balance and XRCD experiments. The black dot is the value recommended by the CODATA in 2017 [42,43] and was adopted by the 26th CGPM to redefine the kilogram.




Figure 2. Measured values of the Planck constant, h , used by CODATA 2017 special adjustment [43]. The numbers shown for the abbreviations denote the year, e.g., “17” stands for 2017. See text and Abbreviations for more details.

The definition of the kilogram does not provide any practical way to realize it. To address this, the CCM issued a *mise en pratique* [41], which explains how the definition can be realized. However, any procedure that derives the mass value in a way that can be traced back to the set of defined fundamental constants, within the stated uncertainty, can be considered a realization of the kilogram. Currently, only the XRCD and the KB experiments can realize the kilogram with relative uncertainties of a few parts in 10^8 , corresponding to a few tens of micrograms.

3.1. The XRCD Method

The XRCD method for the realization of the unit of mass is the reverse of the original experiment determining the Avogadro constant in terms of the kilogram prototype mass, where the measurement equation is solved for the mass of a perfect and hyper-pure Si crystal shaped as a nearly perfect sphere instead of for the Avogadro constant. The link to the Planck constant originates from the independent determination of the mass of the ^{28}Si atom, which is now needed.

According to this method, the measurement equation of the sphere mass is

$$m_{\text{sphere}} = Nm(\text{Si}) - m_{\text{deficit}} + m_{\text{SL}} = \frac{8V_{\text{core}}}{a^3}m(\text{Si}) - m_{\text{deficit}} + m_{\text{SL}}, \quad (3)$$

where N is the number of Si atoms in the Si core of the sphere, $m(\text{Si})$ is the mass of the Si atom, m_{deficit} is the mass defect or excess due to chemical impurities and crystal point defects (self-interstitials and vacancies), m_{SL} is the mass of the surface layers (see Figure 3), a^3 is the (cubic) unit-cell volume, 8 is the number of Si atoms in the unit cell, and V_{core} is the volume of the Si core.




Figure 3. An X-ray interferometer mirrors in a nearly perfect silicon sphere. By combining optical and X-ray interferometry, the lattice parameter, a , of the unit cell volume was determined at INRiM. The macroscopic silicon sphere volume was determined by using optical interferometry NMIJ and PTB. **Top left:** the sketch of the unit cell. **Top right:** the model of the silicon surface layers. Adapted from Ref. [44].

Since silicon occurs in three stable isotopes, $m(\text{Si})$ is given by

$$m(\text{Si}) = \sum_{k=28}^{30} f(^k\text{Si})m(^k\text{Si}), \tag{4}$$

where $m(^k\text{Si})$ is the mass of the k -th isotope, and the amount-of-substance fractions, $f(^k\text{Si})$, of each isotope (^kSi) has to be measured. Taking the identity of the ratio of the masses of (^kSi) and the electron and the ratio of their relative masses, A_r , into account, the mass of (^kSi) can be obtained from the electron mass, $m(e)$, as

$$m(^k\text{Si}) = \frac{A_r(^k\text{Si})}{A_r(e)}m(e). \tag{5}$$

Ultimately, the electron mass is traced back to the Planck constant via the measurements of the Rydberg and fine-structure constants:

$$m(e) = \frac{2hR_\infty}{c\alpha^2}. \tag{6}$$

Combining Equations (3)–(6), the sphere mass is traced back to the Planck constant:

$$m_{\text{sphere}} = \frac{8V_{\text{core}}}{a^3} \frac{2hR_\infty}{c\alpha^2} \left(\sum_{k=28}^{30} f(^k\text{Si}) \frac{A_r(^k\text{Si})}{A_r(e)} \right) - m_{\text{deficit}} + m_{\text{SL}}. \tag{7}$$

3.2. The Kibble Balance Method

The direct way to trace a mass to the Planck constant is by the Kibble balance. In this device, the mechanical and electrical powers produced by the mass motion in the Earth’s gravitational field and the motion of the supporting coil in a magnetic field are virtually compared. As shown in Figure 4, the comparison is carried out in two steps.




Figure 4. Cont.




Figure 4. Phases of a Kibble balance experiment. In the static mode, the apparatus counterbalances the test mass. The acting force, F , generated by the current flowing in the coil is balanced against the weight, mg , of the test mass. The current, I , flowing in the supporting coil of length, L , is measured in terms of Josephson voltage and quantum Hall resistance. U denotes the voltage across the resistor. In the dynamic mode, the test mass is removed, and the coil is moved at constant velocity, v , in the vertical direction through the magnetic field, B . The induced voltage, E , is measured in terms of Josephson voltage. V (circled) and R denote the voltmeter and resistor, respectively.

In the static or force mode, the balance compares the weight (mg , with g the acceleration of gravity) and the Lorentz force BLI are generated by the interaction of the electrical current (I) flowing in the supporting coil of length L and immersed in the magnetic field, B . Hence, by leaving out a vector notation,

$$mg = BLI. \tag{8}$$

In the dynamic or velocity mode, the coil is moved at constant velocity, v , and the induced electromotive force, E , is measured at its ends. Hence,

$$E = BLv. \tag{9}$$

The geometric term BL appearing in both Equations (8) and (9) cannot be measured with the aimed 10^{-8} fractional accuracy. Therefore, it is eliminated to obtain the so-called watt equation (the name stems from both the mechanical and electrical powers being measured in the watt unit)

$$mgv = EI. \tag{10}$$

The electromotive force is measured via the Josephson effect as $E = n_1(h/2e)f_1$, where n_1 is an integer, e is the elementary charge, and f_1 is the frequency irradiating the device. The current measurement, via the Ohm law, $I = U/R$, is based on both the Josephson effect (hence, $U = n_2(h/2e)f_2$), and the quantum Hall effect (hence, $R = r(h/e^2)$). Here, n_2 is an integer, r is a unitless calibration factor, f_2 is a frequency irradiating the Josephson device, and the ratio, $R_k = h/e^2$, is known as the von Klitzing constant.

After using the above definitions in Equation (10), the h/m ratio reads

$$\frac{h}{m} = \frac{4r}{n_1 n_2} \frac{g\nu}{f_1 f_2} \tag{11}$$

This measurement equation, which was previously solved for h (expressed in terms of the mass of the 1901 Pt-Ir prototype of the kilogram), expresses m in terms of the stipulated values of the Planck constant, speed of light, and frequency splitting between the hyperfine levels of the ground state of the ^{133}Cs atom.

All the quantities on the right-hand side of Equation (11) are measured with uncertainties small enough to give h/m with a relative uncertainty of 1×10^{-8} . However, other sources of uncertainty, such as alignments, unwanted motion, and parasitic forces and torques, must be made harmless and still prevent this uncertainty from being reached. Detailed descriptions of the Kibble-balance experiments can be found in Refs. [29–32,45–51].

4. Dissemination of the Kilogram

On 20 May 2019, the kilogram was redefined to be based on the fixed numerical value of the Planck constant, rather than the international prototype’s mass. Actually, after the redefinition, any National Metrological Institute (NMI) could realize that the kilogram is traceable to the Planck constant.

The values measured by Kibble balances and the XRCD method submitted to CODATA for the 2017 special adjustment were not in perfect agreement. To achieve consistency, a multiplicative expansion factor of 1.7 was applied to the uncertainties of the data shown in Figure 2 [42]. This correction to the uncertainties of the data led to the implementation of a procedure for a smooth and reliable transition between the international prototype and the new SI realizations of the kilogram. Therefore, the CCM recommended [52] coordinated (rather than independent) realizations. The coordination key is a “consensus value” for the mass of the IPK resulting from international key comparisons of the realizations of the new kilogram definition. The consensus value is an offset from the BIPM as-maintained mass unit, which represents the mass of the IPK. It acts as a proxy for the realization experiments, and its uncertainty reflects a typical uncertainty of the experiments participating in the comparisons and the stability of the BIPM working standards.

The details were established by a CCM task group [53] as follows:

- Phase 0: traceability to the IPK, $m_{\text{IPK}} = 1 \text{ kg}$, before the revision of the SI on 20 May 2019;
- Phase 1: traceability to the Planck constant via its known relationship with the IPK, $m_{\text{IPK}} = 1 \text{ kg}$, with an additional uncertainty of $u_{m_{\text{IPK}}} = 10 \text{ }\mu\text{g}$, from 20 May 2019 until a consensus value resulting from the first key comparison of primary realizations of the kilogram is published.
- Phase 2: dissemination from the consensus value, until the decision of the CCM.
- Phase 3: dissemination by individual realizations.

Three comparison exercises have been already performed: a pilot study [54] and two international key comparisons [55,56]. Table 1 gives the consensus values of the mass of the international prototype following these exercises. The 2023 consensus value for the SI unit of mass, the kilogram, has been determined to be $1 \text{ kg} - 7 \text{ }\mu\text{g}$, with a standard uncertainty of $20 \text{ }\mu\text{g}$ to the mass value of the IPK, which is equal to the BIPM as-maintained mass unit. Since 1 March 2023, to achieve consistency with the consensus value, all NMIs would need to reduce the mass value of their national as-maintained mass unit by $7 \text{ }\mu\text{g}$ with respect to the mass value based on the IPK.

Table 1. Implementation of the traceability of the kilogram over the years 2019, 2021, and 2023 [54–56].

Date of Implementation	Basis for Dissemination	Uncertainty
20 May 2019	$m(\text{IPK}) = 1 \text{ kg}$	$10 \text{ }\mu\text{g}$
1 February 2021	Consensus value 2021 $m(\text{IPK}) = 1 \text{ kg} - 2 \text{ }\mu\text{g}$	$20 \text{ }\mu\text{g}$
1 March 2023	Consensus value 2023 $m(\text{IPK}) = 1 \text{ kg} - 7 \text{ }\mu\text{g}$	$20 \text{ }\mu\text{g}$

It should be noted that, during the last key comparison, the difference between the two primary realizations with the lowest uncertainty using the Kibble balance (at the NRC) and the XRCD method (at the PTB) is 0.050 mg, which is about five times the stated uncertainty (coverage factor $k = 1$) of the individual experiments [56].

Considering the current uncertainty of the 2023 consensus value (which includes all primary realizations participating in the last three comparison exercises), it is possible that the Kibble balance and the count of atoms in the ^{28}Si realizations could detect any instability in the international prototype within a few decades.

5. Conclusions

This review focused on how the International System of Units (SI) has been redefined by establishing its units on stipulated values of a set of fundamental constants of physics. Redefining the measurement units required new technologies to realize them in practice and without introducing discontinuity of their value. Since 1990, the redefinition of electric units and the mole would have been possible based on the Planck and Avogadro constants and the elementary charge. Such redefinition would have implied a fixed value of the mass difference between the ground-state hyperfine energy levels of the ^{133}Cs atom, which, in turn, would have necessitated a redefinition of the unit of mass. However, there was no technology available to derive a kilogram from this very small mass difference in a manner competitive with the dissemination from the international prototype.

Hence, the central role played by the development of the technologies necessary to reliably realize the kilogram by reversing accurate measurements of the Planck and Avogadro constants (the pillars of today's SI) was carried out by comparing electrical and mechanical powers and counting atoms. Thanks to the accuracy required, approaching 10^{-8} in fractional terms, these researches and developments have been quite a challenge [57] and, then, were at the top of the metrologist's agenda for many decades; and engaged a multitude of those involved.

In the new system, any unit can be expressed by the fixed values of one or more of the seven defining constants. The links between the units historically identified as "base units" and the defining constants are displayed by the SI logo shown in Figure 5. However, the distinction between the second, meter, kilogram, ampere, kelvin, mole, and candela, conventionally chosen as base units, and the other derived units, apart from historical and educational values, no longer has any reason to exist.




Figure 5. The logo of the International System of Units (SI).

In the new SI, any measurement procedure (no matter what its accuracy is) giving a result without relation to a standard, but relating it directly to defining constants, is a primary realization of the unit (or its multiples and submultiples) within the stated uncertainty. This answers the “for all times, for all peoples” request of the French revolutionaries and opens the way to realize, to mention a few, the mechanical units of mass, force, and torque—ranging from atomic to macroscopic scales—where needed, tailored on the actual accuracy needs, and, since independent of material artefacts, traceable to the system of units independently of national metrology institutes.

From this vantage, the KB and XRCD method are both primary realizations of the kilogram and, possibly, its multiples and submultiples. It is fortunate there exist independent, cross-checking, and top-level realizations relying on different aspects of nature’s developments. While the KB relies on solid-state physics (via the Josephson and quantum Hall effects), the XRCD realization ultimately relies on the measurement of the ratios between atoms or electron inertial masses and the Planck constant (via atomic interferometry and spectroscopy).

Practical considerations must also be taken into account. Through silicon crystallization, the XRCD method amplifies an atomic mass to macroscopic scales and allows material standards to be realized. This means that, once the realization has been completed, the well-established mass metrology can be unaffected by the redefinition. However, if one aims at achieving the smallest realization uncertainty at the one-kilogram level, isotopic enrichment is necessary and, consequently, the primary realization cost is exceptionally high given the (present) impossibility of determining the isotopic composition of natural silicon with the necessary accuracy.

The cost of building a KB with a comparable level of accuracy is less expensive and more widely affordable. However, the operation cost can be quite high. Dealing with the KB is a delicate matter, and keeping it operational requires a team of highly skilled and versatile professionals.

Funding: This research received no external funding.

Data Availability Statement: No new data were created or analyzed in this study. Data sharing is not applicable to this article.

Conflicts of Interest: The author declares no conflict of interest.

Abbreviations

The following abbreviations are used in this manuscript:

BIPM	Bureau International des Poids et Mesures (International Bureau of Weights and Measures)
CGPM	Conférence Générale des Poids et Mesures (General Conference on Weights and Measures)
CIPM	Comité International des Poids et Mesures (International Committee for Weights and Measures)
CCM	Consultative Committee for Mass and Related Quantities
CCU	Consultative Committee for Units
CODATA	Committee on Data for Science and Technology
IAC	International Avogadro Coordination
INRiM	Istituto Nazionale di Ricerca Metrologica
IRMM	Institute for Reference Material and Measurements—European Commission Joint Research Center
IPK	International Prototype Kilogram
KB	Kibble balance
LNE	Laboratoire National de Métrologie et d’Essais

LSA	Least-squares adjustment
MRA	Mutual Recognition Arrangement
NIST	National Institute of Standards and Technology
NMI	National Metrology Institute
NMIA	National Measurement Institute—Australia
NMIJ	National Measurement Institute of Japan
NPL	National Physical Laboratory
NRC	National Research Council (Canada)
PTB	Physikalisch-Technische Bundesanstalt
SI	Système International d’unités (International System of Units)
XRCD	X-ray Crystal Density

References

- Wallis Budge, E.A. The Egyptian Book of the Dead. Available online: <https://bookofthedead.book-lover.com> (accessed on 25 April 2024).
- Documents Diplomatiques de la Conférence du Mètre. *Convention du Mètre*; Imprimerie Nationale: Paris, France, 1875. Available online: <https://www.bipm.org/en/founding-documents> (accessed on 25 April 2024).
- Smith, F.J. Standard Kilogram Weights: A Story of Precision Fabrication. *Platin. Met. Rev.* **1973**, *17*, 66–68. [[CrossRef](#)]
- Davis, R. The SI Unit of Mass. *Metrologia* **2003**, *40*, 299–305. [[CrossRef](#)]
- Comptes Rendus des Séances de la Première Conférence Générale des Poids et Mesures, réunie à Paris en 1889*; Gauthier-Villars et Fils, Imprimeurs-Libraires de l’École Polytechnique, du Bureau des Longitudes: Paris, France, 1890; p. 38. [[CrossRef](#)]
- Comptes Rendus des Séances de la Troisième Conférence Générale des Poids et Mesures, réunie à Paris en 1901*; Gauthier-Villars et Fils, Imprimeurs-Libraires de l’École Polytechnique, du Bureau des Longitudes: Paris, France, 1901; p. 70. [[CrossRef](#)]
- Girard, G. The Third Periodic Verification of National Prototypes of the Kilogram (1988–1992). *Metrologia* **1994**, *31*, 317–336. [[CrossRef](#)]
- Stock, M.; Barat, P.; Davis, R.S.; Picard, A.; Milton, M.J.T. Calibration Campaign against the International Prototype of the Kilogram in Anticipation of the Redefinition of the Kilogram Part I: Comparison of the International Prototype with Its Official Copies. *Metrologia* **2015**, *52*, 310–316. [[CrossRef](#)]
- Maxwell, J.C. Address by the the President of the Section. In *Report of the Fortieth Meeting of the British Association for the Advancement Science Held at Liverpool in September 1870*; John Murray: London, UK, 1871; Notices and Abstracts of Miscellaneous Communications to the Sections; pp. 1–9. Available online: <https://www.biodiversitylibrary.org/bibliography/2276> (accessed on 25 April 2024).
- Planck, M. Über Irreversible Strahlungsvorgänge. *Ann. Phys.* **1900**, *306*, 69–122. [[CrossRef](#)]
- Egidi, C. Phantasies on a Natural Unity of Mass. *Nature* **1963**, *200*, 61–62. [[CrossRef](#)]
- Bonse, U.; Hart, M. An X-ray Interferometer. *Appl. Phys. Lett.* **1965**, *6*, 155–156. [[CrossRef](#)]
- Deslattes, R.D.; Henins, A.; Bowman, H.A.; Schoonover, R.M.; Carroll, C.L.; Barnes, I.L.; Machlan, L.A.; Moore, L.J.; Shields, W.R. Determination of the Avogadro Constant. *Phys. Rev. Lett.* **1974**, *33*, 463–466. [[CrossRef](#)]
- Fujii, K.; Waseda, A.; Kuramoto, N.; Mizushima, S.; Becker, P.; Bettin, H.; Nicolaus, A.; Kuetgens, U.; Valkiers, S.; Taylor, P.; et al. Present State of the Avogadro Constant Determination From Silicon Crystals With Natural Isotopic Compositions. *IEEE Trans. Instrum. Meas.* **2005**, *54*, 854–859. [[CrossRef](#)]
- Zosi, G. A Neo-Pythagorean Approach towards an Atomic Mass Standard. *Lett. Nuovo Cim.* **1983**, *38*, 577–580. [[CrossRef](#)]
- Leistner, A.; Zosi, G. Polishing a 1-kg Silicon Sphere for a Density Standard. *Appl. Opt.* **1987**, *26*, 600. [[CrossRef](#)] [[PubMed](#)]
- Andreas, B.; Azuma, Y.; Bartl, G.; Becker, P.; Bettin, H.; Borys, M.; Busch, I.; Gray, M.; Fuchs, P.; Fujii, K.; et al. Determination of the Avogadro Constant by Counting the Atoms in a ²⁸Si Crystal. *Phys. Rev. Lett.* **2011**, *106*, 030801. [[CrossRef](#)] [[PubMed](#)]
- Andreas, B.; Azuma, Y.; Bartl, G.; Becker, P.; Bettin, H.; Borys, M.; Busch, I.; Fuchs, P.; Fujii, K.; Fujimoto, H.; et al. Counting the Atoms in a ²⁸Si Crystal for a New Kilogram Definition. *Metrologia* **2011**, *48*, S1–S13. [[CrossRef](#)]
- Azuma, Y.; Barat, P.; Bartl, G.; Bettin, H.; Borys, M.; Busch, I.; Cibik, L.; D’Agostino, G.; Fujii, K.; Fujimoto, H.; et al. Improved Measurement Results for the Avogadro Constant Using a ²⁸Si-enriched Crystal. *Metrologia* **2015**, *52*, 360–375. [[CrossRef](#)]
- Bartl, G.; Becker, P.; Beckhoff, B.; Bettin, H.; Beyer, E.; Borys, M.; Busch, I.; Cibik, L.; D’Agostino, G.; Darlatt, E.; et al. A New ²⁸Si Single Crystal: Counting the Atoms for the New Kilogram Definition. *Metrologia* **2017**, *54*, 693–715. [[CrossRef](#)]
- Fujii, K.; Massa, E.; Bettin, H.; Kuramoto, N.; Mana, G. Avogadro Constant Measurements Using Enriched ²⁸Si Monocrystals. *Metrologia* **2018**, *55*, L1–L4. [[CrossRef](#)]
- Basile, G.; Bergamin, A.; Cavagnero, G.; Mana, G.; Vittone, E.; Zosi, G. The (220) Lattice Spacing of Silicon. *IEEE Trans. Instrum. Meas.* **1995**, *44*, 526–529. [[CrossRef](#)]
- Massa, E.; Sasso, C.P.; Mana, G.; Palmisano, C. A More Accurate Measurement of the ²⁸Si Lattice Parameter. *J. Phys. Chem. Ref. Data* **2015**, *44*, 031208. [[CrossRef](#)]

24. Bartl, G.; Bettin, H.; Krystek, M.; Mai, T.; Nicolaus, A.; Peter, A. Volume Determination of the Avogadro Spheres of Highly Enriched ^{28}Si with a Spherical Fizeau Interferometer. *Metrologia* **2011**, *48*, S96–S103. [CrossRef]
25. Kuramoto, N.; Azuma, Y.; Inaba, H.; Fujii, K. Volume Measurements of ^{28}Si -enriched Spheres Using an Improved Optical Interferometer for the Determination of the Avogadro Constant. *Metrologia* **2017**, *54*, 193–203. [CrossRef]
26. Pramann, A.; Rienitz, O.; Schiel, D.; Schlote, J.; Güttler, B.; Valkiers, S. Molar Mass of Silicon Highly Enriched in ^{28}Si Determined by IDMS. *Metrologia* **2011**, *48*, S20–S25. [CrossRef]
27. Massa, E.; Nicolaus, A. International Determination of the Avogadro Constant. *Metrologia* **2011**, *48*, E01. [CrossRef]
28. Kibble, B.P. A Measurement of the Gyromagnetic Ratio of the Proton by the Strong Field Method. In *Atomic Masses and Fundamental Constants 5*; Sanders, J.H., Wapstra, A.H., Eds.; Springer Science+Business Media: New York, NY, USA, 1976; pp. 545–551. [CrossRef]
29. Robinson, I.A. Towards the Redefinition of the Kilogram: A Measurement of the Planck Constant Using the NPL Mark II Watt Balance. *Metrologia* **2012**, *49*, 113–156. [CrossRef]
30. Schlamminger, S.; Haddad, D.; Seifert, F.; Chao, L.S.; Newell, D.B.; Liu, R.; Steiner, R.L.; Pratt, J.R. Determination of the Planck Constant Using a Watt Balance with a Superconducting Magnet System at the National Institute of Standards and Technology. *Metrologia* **2014**, *51*, S15–S24. [CrossRef]
31. Sanchez, C.A.; Wood, B.M.; Green, R.G.; Liard, J.O.; Inglis, D. A Determination of Planck's Constant Using the NRC Watt Balance. *Metrologia* **2014**, *51*, S5–S14. [CrossRef]
32. Thomas, M.; Ziane, D.; Pinot, P.; Karcher, R.; Imanaliev, A.; Dos Santos, F.P.; Merlet, S.; Piquemal, F.; Espel, P. A Determination of the Planck Constant Using the LNE Kibble Balance in Air. *Metrologia* **2017**, *54*, 468–480. [CrossRef]
33. Tiesinga, E.; Mohr, P.J.; Newell, D.B.; Taylor, B.N. CODATA Recommended Values of the Fundamental Physical Constants: 2018. *Rev. Mod. Phys.* **2021**, *93*, 025010. [CrossRef] [PubMed]
34. Mills, I.M.; Mohr, P.J.; Quinn, T.J.; Taylor, B.N.; Williams, E.R. Redefinition of the Kilogram: A Decision Whose Time Has Come. *Metrologia* **2005**, *42*, 71–80. [CrossRef]
35. *Proceedings of the 21st General Conference on Weights and Measures, 11–15 October 1999*; BIPM: Sèvres, France, 2001; p. 331. [CrossRef]
36. *Proceedings of the 21st General Conference on Weights and Measures, November 2007*; BIPM: Sèvres, France, 2010; p. 434. [CrossRef]
37. *Proceedings of the 24th General Conference on Weights and Measures, October 2011*; BIPM: Sèvres, France, 2013; p. 532. [CrossRef]
38. *Proceedings of the 25th General Conference on Weights and Measures, November 2014*; BIPM: Sèvres, France, 2015; p. 416. [CrossRef]
39. Richard, P.; Fang, H.; Davis, R. Foundation for the Redefinition of the Kilogram. *Metrologia* **2016**, *53*, A6–A11. [CrossRef]
40. Consultative Committee for Mass and Related Quantities (CCM). *Report of the 14th Meeting (21–22 February 2013) to the International Committee for Weights and Measures*; BIPM: Sèvres, France, 2013; p. 36. [CrossRef]
41. BIPM. Practical Realization of the Definition of Some Important Units. Available online: <https://www.bipm.org/en/publications/mises-en-pratique> (accessed on 25 April 2024).
42. Newell, D.B.; Cabiati, F.; Fischer, J.; Fujii, K.; Karshenboim, S.G.; Margolis, H.S.; De Mirandés, E.; Mohr, P.J.; Nez, F.; Pachucki, K.; et al. The CODATA 2017 Values of h , e , k , and N_A for the Revision of the SI. *Metrologia* **2018**, *55*, L13–L16. [CrossRef]
43. Mohr, P.J.; Newell, D.B.; Taylor, B.N.; Tiesinga, E. Data and Analysis for the CODATA 2017 Special Fundamental Constants Adjustment. *Metrologia* **2018**, *55*, 125–146. [CrossRef]
44. Fujii, K.; Bettin, H.; Becker, P.; Massa, E.; Rienitz, O.; Pramann, A.; Nicolaus, A.; Kuramoto, N.; Busch, I.; Borys, M. Realization of the kilogram by the XRCD method. *Metrologia* **2016**, *53*, A19–A45. [CrossRef]
45. Robinson, I.A. Watt and Joule Balances. *Metrologia* **2014**, *51*, S1–S3. [CrossRef]
46. Fang, H.; Bielsa, F.; Li, S.; Kiss, A.; Stock, M. The BIPM Kibble Balance for Realizing the Kilogram Definition. *Metrologia* **2020**, *57*, 045009. [CrossRef]
47. Eichenberger, A.; Baumann, H.; Mortara, A.; Tommasini, D.; Reber, D.; Klingelé, E.; Jeanneret, B.; Jeckelmann, B. First Realisation of the Kilogram with the METAS Kibble Balance. *Metrologia* **2022**, *59*, 025008. [CrossRef]
48. Wood, B.M.; Sanchez, C.A.; Green, R.G.; Liard, J.O. A Summary of the Planck Constant Determinations Using the NRC Kibble Balance. *Metrologia* **2017**, *54*, 399–409. [CrossRef]
49. Haddad, D.; Seifert, F.; Chao, L.S.; Possolo, A.; Newell, D.B.; Pratt, J.R.; Williams, C.J.; Schlamminger, S. Measurement of the Planck Constant at the National Institute of Standards and Technology from 2015 to 2017. *Metrologia* **2017**, *54*, 633–641. [CrossRef]
50. Kim, D.; Kim, M.; Seo, M.; Woo, B.-C.; Lee, S.; Kim, J.-A.; Chae, D.-H.; Kim, M.-S.; Choi, I.-M.; Lee, K.-C. Realization of the Kilogram Using the KRIS Kibble Balance. *Metrologia* **2020**, *57*, 055006. [CrossRef]
51. Ahmedov, H.; Orhan, R.; Korutlu, B. UME Kibble Balance Operating in Air. *Metrologia* **2023**, *60*, 015003. [CrossRef]
52. Recommendation 1 of the CCM (2017). Available online: <https://www.bipm.org/en/-/ccm-2017-recommendation-1> (accessed on 25 April 2024).
53. CCM/2019-06B: CCM Detailed Note on the Dissemination Process after the Redefinition of the Kilogram. Available online: <https://www.bipm.org/en/committees/cc/ccm/17-2019> (accessed on 25 April 2024).
54. BIPM. Final Reports: Pilot Studies-M. Available online: <https://www.bipm.org/en/committees/cc/ccm/pilot-studies> (accessed on 25 April 2024).
55. Stock, M.; Conceição, P.; Fang, H.; Bielsa, F.; Kiss, A.; Nielsen, L.; Kim, D.; Kim, M.; Lee, K.-C.; Lee, S.; et al. Report on the CCM Key Comparison of Kilogram Realizations CCM.M-K8.2019. *Metrologia* **2020**, *57*, 07030. [CrossRef]

-
56. Stock, M.; Conceição, P.; Fang, H.; Bielsa, F.; Kiss, A.; Nielsen, L.; Beaudoux, F.; Espel, P.; Thomas, M.; Ziane, D.; et al. Final Report on the CCM Key Comparison of Kilogram Realizations CCM.M-K8.2021. *Metrologia* **2023**, *60*, 07003. [[CrossRef](#)]
 57. Jones, N. Frontier Experiments: Tough Science. *Nature* **2012**, *481*, 14–17. [[CrossRef](#)] [[PubMed](#)]

Disclaimer/Publisher’s Note: The statements, opinions and data contained in all publications are solely those of the individual author(s) and contributor(s) and not of MDPI and/or the editor(s). MDPI and/or the editor(s) disclaim responsibility for any injury to people or property resulting from any ideas, methods, instructions or products referred to in the content.