

Article

Modeling and Design of a Spherical Remote Center-of-Motion Surgical Robot [†]

Calin Vaida ^{1,2}, Daniel Horvath ^{1,2}, Ionut Zima ^{1,2}, Marius Miclaus ^{1,2}, Bogdan Gherman ^{1,2}, Corina Radu ³, Paul Tucan ^{1,2}, Stefan Vegh ^{1,2}, Dragos Sebeni ^{1,2}, Adrian Pislă ^{1,2}, Damien Chablat ^{1,2,4}, Nadim Al Hajjar ⁵ and Doina Pislă ^{1,2,6,*}

¹ Research Centre for Robots Simulation and Testing, Technical University of Cluj-Napoca, Memorandumului 28, 400114 Cluj-Napoca, Romania; calin.vaida@mep.utcluj.ro (C.V.); horvath.daniel@mep.utcluj.ro (D.H.); ionut.zima@mep.utcluj.ro (I.Z.); marius.miclaus@mep.utcluj.ro (M.M.); bogdan.gherman@mep.utcluj.ro (B.G.); paul.tucan@mep.utcluj.ro (P.T.); stefan.vegh@mep.utcluj.ro (S.V.); dragos.sebeni@mep.utcluj.ro (D.S.); adrian.pisla@muri.utcluj.ro (A.P.); damien.chablat@cnrs.fr (D.C.)

² European University of Technology, European Union

³ Department of Internal Medicine, “Iuliu Hatieganu” University of Medicine and Pharmacy, 400347 Cluj-Napoca, Romania; corina.radu@umfcluj.ro

⁴ École Centrale Nantes, Nantes Université, CNRS, LS2N, UMR 6004, F-44000 Nantes, France

⁵ Department of Surgery, “Iuliu Hatieganu” University of Medicine and Pharmacy, 400347 Cluj-Napoca, Romania; nadim.alhajjar@umfcluj.ro

⁶ Technical Sciences Academy of Romania, B-dul Dacia, 26, 030167 Bucharest, Romania

* Correspondence: doina.pisla@mep.utcluj.ro

[†] This paper is an extended version of our paper published in the 3rd International Conference on Advanced Autonomous Systems, Bucharest, Romania, 18–20 November 2026.

Abstract

Remote center-of-motion mechanisms are essential in minimally invasive surgery because they allow surgical instruments or an endoscopic camera to pivot around a trocar entry point while eliminating lateral motion at the incision. This paper presents the design, kinematic modeling, prototype implementation and preliminary evaluation under laboratory conditions of a compact, spherical, remote center-of-motion robot for minimally invasive surgical orientation tasks. The proposed mechanism uses a spherical kinematic architecture actuated by a contra-rotating differential gearbox. This gearbox generates two coaxial output rotations of equal magnitude and opposite direction from a single input, mechanically synchronizing the opposed motion of the two base links and eliminating the need for cable-pulley transmission or dual electronically synchronized motors. A second actuator chain rotates the gearbox assembly around the base axis, thereby decoupling the extension–retraction motion from base-axis rotation. Forward and inverse kinematic formulations were derived for teleoperation of the robot using a 7 degrees of freedom haptic device and for remote center-of-motion orientation control using a 3-axis joystick. A proof-of-concept prototype was developed and integrated with a custom embedded controller, closed-loop motor control, a master-console interface and video feedback loop. The system was evaluated in a phantom-torso setup using a custom endoscopic camera, internal visual markers and an OptiTrack-based measurement of the remote center-of-motion accuracy. The qualitative experiment confirmed functional integration of the mechanical, electronic and software subsystems, while the optical-tracking measurement showed that the pivot constraint was maintained with a mean deviation of 1.69 mm and a root-mean-square deviation of 2.13 mm over the analyzed orientation sweep. The main limitations remain the 1:1 gearbox ratio, limited actuator torque, additively manufactured

Academic Editor: Tamás Haidegger

Received: 4 June 2026

Revised: 7 July 2026

Accepted: 9 July 2026

Published: 17 July 2026

Copyright: © 2026 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the [Creative Commons Attribution \(CC BY\)](https://creativecommons.org/licenses/by/4.0/) license.

gearing and the absence of repeated-trial repeatability and full workspace characterization.

Keywords: surgical robot; remote center of motion; kinematics; modeling; minimally invasive surgery; spherical mechanism; design

1. Introduction

Cancer remains one of the leading causes of mortality worldwide, with approximately 20 million new cases reported in 2022, and incidence rates expected to continue increasing by 2040 [1,2]. Among all cancer types, esophageal cancer ranks among the most aggressive malignancies, being the eleventh most common cancer and one of the leading causes of cancer-related death globally [3,4]. Due to the absence of specific symptoms in its early stages, many patients are diagnosed only after the disease has advanced, resulting in a poor 5-year survival rate of approximately 15–25%.

Esophageal cancer is broadly classified into two major histological types: esophageal squamous cell carcinoma (ESCC) and esophageal adenocarcinoma (EAC). ESCC is the predominant form worldwide and is associated primarily with smoking, alcohol consumption, poor nutrition, thermal injury from hot beverages and human papillomavirus infection [5]. In contrast, EAC is more common in Western countries and is closely associated with Barrett's esophagus, gastroesophageal reflux disease (GERD), obesity and smoking [6]. While modifiable risk factors include smoking, alcohol use, obesity and dietary habits, non-modifiable factors involve age, genetics and family history.

Current treatment strategies for esophageal cancer depend on tumor stage and patient condition and typically include combinations of chemotherapy, radiotherapy, immunotherapy and surgery. Among these, esophagectomy remains the primary curative treatment for resectable tumors. The procedure involves partial or complete removal of the esophagus together with mediastinal lymph nodes. Although surgical outcomes have improved over recent decades, esophagectomy remains one of the most technically demanding procedures due to the anatomical location of the esophagus behind major thoracic organs and its extension across the neck, thorax and abdomen [7]. Common surgical approaches include the transthoracic Ivor Lewis and McKeown procedures, as well as the transhiatal approach [8]. Despite advances in perioperative care, esophagectomy is still associated with considerable postoperative morbidity and mortality [6].

The increasing integration of robotic systems into surgical rooms has significantly influenced minimally invasive surgery. Since the approval of the Intuitive Surgical da Vinci Surgical System in 2000, robotic-assisted surgery has demonstrated benefits such as reduced blood loss, improved precision, lower postoperative pain and shorter hospital stays [9]. Most robotic procedures are currently concentrated in urology, gynecology and general surgery, although the adoption of robotic systems continues to expand [9,10]. The main advantages of these systems include enhanced dexterity, three-dimensional visualization, tremor filtering and improved surgeon ergonomics.

Beyond mechanical design and teleoperation, artificial intelligence is increasingly used to move surgical robots toward supervised autonomy [11,12]. For a proof-of-concept spherical RCM mechanism such as the one proposed here, its most immediate role is the compensation of known hardware imperfections rather than task-level autonomy. Because the prototype uses a 1:1 differential ratio and additively manufactured gearing, its trajectory tracking and RCM-holding accuracy are limited by backlash, joint friction and coarse motion resolution. These effects are difficult to model analytically but are well suited to data-driven compensation: adaptive neuro-fuzzy inference systems (ANFIS) or

reinforcement-learning-based controller auto-tuning can learn the relationship between commanded and actual joint motion and adapt the control gains online, improving trajectory tracking and reducing the residual RCM deviation without redesigning the mechanism [13]. The encoder and motor-current signals already acquired by the embedded controller (Section 3.3) additionally allow abnormal operating conditions to be detected as a supervisory safety layer, while the surgeon remains in full control. These compensation directions are revisited in Section 5.6.

In recent years, multiple robotic platforms have emerged alongside the da Vinci system, including Senhance [14], Hugo RAS [15], Versius [16,17] and Hinotori [18]. These systems differ mainly in the implementation of the remote center of motion (RCM), a critical requirement in minimally invasive surgery that ensures instrument motion pivots around the trocar insertion point. Mechanical RCM architectures enforce this constraint through linkage geometry, while software-based solutions rely on redundant manipulators and controller integrity. Mechanical RCM systems provide inherent safety and independence from sensor calibration, whereas software-constrained systems offer greater flexibility, including real-time relocation of the RCM point [19].

Motivated by these challenges, the present work introduces a compact and scalable spherical RCM mechanism intended for minimally invasive surgical applications. The proposed design operates similarly to a parallelogram-based mechanism while introducing a novel contra-rotating gearbox that removes the need for cable and pulley transmission systems. The mechanism is designed both as a surgical instrument positioner and as part of a future multi-manipulator robotic platform capable of teleoperated control through a master console [20,21].

The main contributions of this work are: (i) the development of a compact spherical RCM robot for minimally invasive orientation tasks; (ii) the design of a contra-rotating differential gearbox enabling synchronized motion and decoupled extension–retraction control; (iii) the derivation of forward and inverse kinematic models for teleoperation using either a 7-DoF haptic interface or a 3-axis joystick; (iv) the implementation of a complete proof-of-concept robotic platform integrating embedded control, motor actuation and endoscopic feedback; and (v) the qualitative and quantitative evaluation of the system in a phantom-torso experimental setup, together with the identification of limitations that must be addressed before quantitative preclinical evaluation.

2. State of the Art on Remote Center-of-Motion Surgical Robotic Systems

As discussed in the previous section, robot-assisted surgery delivers many advantages in terms of positional accuracy and stability, and much of that success is owed to the mechanical architecture of the system. Remote center-of-motion (RCM) mechanisms have played a fundamental role in surgical robots, especially in minimally invasive procedures, due to their ability to constrain the movement of surgical instruments to precisely pivot around a fixed point in space, while also allowing various other motions such as insertion, pitch and yaw. Many designs have been explored over the decades, with some of the more successful designs already being used in surgeries around the world in robots such as the da Vinci systems, which use cable-driven double-parallelogram structures. Almost all robotic platforms that have been and are being used in minimally invasive esophagectomy (MIE) incorporate some form of control around an RCM point. Such mechanisms are undoubtedly advantageous and inherently safe, as they remove the risk of tissue damage caused by human error due to fatigue or hand tremors, and enable a greater degree of precision.

2.1. Mechanically Constrained RCM Architectures

One of the most popular examples of mechanically constrained RCM architectures can be found in the da Vinci system's actuated arms [22], which are based on a parallelogram-type mechanism and use a 3-DoF manipulator to control the orientation around the RCM of an articulated end-effector that also provides an additional 3-DoF [23]. The same family of architectures underlies several other commercial systems, including the LARS surgical assistant developed at Johns Hopkins University and IBM [24], the JHU MINI-RCM module of Stoianovici et al. [25], and the Senhance [14] and Meere Revo-i [26] systems, which share the planar parallelogram lineage.

A similar design architecture was presented by Ming Bai et al. [23] in a robot developed for minimally invasive surgery, based on a classical rigid-linkage architecture that physically enforces the pivoting point at the trocar. Despite this, the authors also integrated a hybrid error compensation approach, combining kinematic modeling with machine learning corrections, to combat imperfections stemming from the mechanical design itself. This approach improves positional fidelity and reduces cumulative errors at the RCM and tool center point without requiring extra sensors or major hardware modifications.

An RCM-based nasal endoscope holder was developed and presented by Xinxin Dai et al. [27] with the purpose of serving as the surgeon's "third hand", effective in reducing fatigue during procedures where multiple instruments need to be handled. Similarly to the da Vinci architecture, this system utilizes a belt-driven parallelogram mechanism with coupled rotational linkages and provides two rotational degrees of freedom around the RCM with an additional axial insertion.

Kai Guo et al. [28] also developed a robotic structure for minimally invasive puncture surgery, with a constrained RCM mechanism based on a modified double-parallelogram linkage, augmented with a telescoping component for axial insertion of the needle. This design achieves two DoF around the fixed pivot point, as well as inherent axial insertion without the need for additional prismatic actuators.

2.2. Software-Constrained RCM on Redundant Manipulators

A representative example of software-constrained RCM control was presented by Kastritsi and Doulgeri [29], who implemented a joint-space admittance controller on a 7-DoF KUKA LWR4+ (KUKA AG, Augsburg, Germany) collaborative robot for surgical instrument manipulation. Their method uses the robot's joint torque sensing and kinematic model to estimate external interaction forces and to generate compliant motion while preserving the trocar constraint. The RCM condition is enforced by computing the deviation between the instrument shaft and the desired pivot point and projecting the commanded joint velocities into a constraint-consistent subspace. This approach illustrates the flexibility of software-imposed RCM strategies, but also their dependence on accurate sensing, calibration and real-time control.

Another possible example of a robotic platform that implements RCM control at a virtual, kinematic software-algorithm level is the Versius (CMR Surgical, Cambridge, UK) surgical system from CMR Surgical [16,17], which utilizes 7-DoF manipulators that feature several uniaxial joints with wristed instrument tips to provide additional articulation. Similar software-RCM strategies are used in the Hugo RAS [15], the Hinotori [18], the Avatera [30,31] and the MiroSurge system (German Aerospace Center/DLR, Oberpfaffenhofen, Germany) [32], illustrating the broader trend in recent commercial platforms toward software-imposed pivot point constraints on redundant serial manipulators.

Nasiri et al. [33] proposed an admittance-based RCM control strategy for a 7-DoF manipulator equipped with a custom surgical instrument module and teleoperated through a haptic device. Their system combines master-slave teleoperation, force/torque

sensing and redundancy resolution to adaptively maintain instrument alignment at the trocar during manipulation. This work illustrates the potential of software-constrained RCM control for adaptive trocar alignment, but also highlights its dependence on force sensing, calibration and real-time control performance [33].

2.3. Spherical RCM Mechanisms

El Jjouaoui et al. [34] investigated a tensegrity-based spherical RCM mechanism composed of two spherical structures with different radii connected through the base and end-effector frames. The system provides two rotational DoF around the RCM point and is actuated through cable-pulley transmissions that control extension and retraction by maintaining coordinated tension in the mechanism. Although this architecture demonstrates the feasibility of compact spherical RCM motion, its reliance on cable actuation introduces additional requirements related to tension management, synchronization and mechanical complexity.

2.4. Prior Art on Differential and Contra-Rotating Gear Actuation in Surgical Robotics

Table 1 summarizes main categories of RCM mechanisms together with representative examples and a qualitative comparison of advantages and limitations relevant to the proposed contra-rotating gearbox-driven spherical RCM.

Table 1. Principal families of RCM mechanisms for surgical robotic applications.

Mechanism Family	Representative Examples	Key Advantages	Principal Limitations
Parallelogram (mechanical)	Da Vinci [22]; Senhance [14]; Revo-i [26]; LARS [24]; JHU MINI-RCM [25]; Dai et al. [27]; Guo et al. [28]; Bai et al. [23]	Architecturally deterministic RCM; clinically mature; well-characterized quantitative performance	Non-negligible planar footprint; cable/belt pretension management; complex multi-arm setups
Spherical, cable-driven	Pantoscope [35]; Raven-II [36]; tensegrity RCM [34]	Compact around RCM; low distal inertia; flexible cable routing	Cable elongation, stiction, periodic re-tensioning; slack detection required
Spherical, dual-motor synchronized	Spherical 5-bar variants [37,38]; Argos [39]	No cables; geometric decoupling possible	Software-level synchronization; double actuator count; vulnerable to motor mismatch
Software-constrained on redundant arms	Hugo RAS [15]; Versius [16,17]; Avatera [30,31]; Hinotori [18]; MiroSurge [32]; Kastritsi & Doulgeri [29]; Nasiri et al. [33]	RCM position adaptable in real time; reconfigurable; uses standard cobots	Constraint depends on controller and sensor integrity; not architecturally safe; calibration-sensitive
Hybrid spherical wrist + cart docking	Distalmotion Dexter [40,41]; Dionis [42]	Compact wrist; clinical hybridization with laparoscopy	Trocar pivot enforced jointly by wrist geometry and docking; complex assembly
Spherical with contra-rotating gearbox (this work)	Proposed mechanism	Single-actuator synchronized opposed base-link motion; mechanically decoupled from base-axis rotation; no cables, no pulleys, no dual motors	Currently 2-DoF demonstrator (no insertion axis); 1:1 differential ratio limit motion resolution; 3D-printed gearing constrains torque

2.5. Recent Commercial Surgical Robotic Platforms

Table 2 provides a complementary overview of selected commercial and research surgical robotic platforms relevant to minimally invasive surgery, with an emphasis on the adopted RCM enforcement strategy, number of DoF and transmission. The comparison positions the proposed prototype relative to clinically established and research-oriented systems, highlighting its mechanically constrained spherical RCM architecture and contra-rotating gearbox actuation.

Table 2. Selected recent commercial and research surgical robotic platforms relevant to MIS, with their RCM-enforcement strategy and transmission type. The last row shows the position of the proposed mechanism.

Platform	First Clearance	DoF/Arm	RCM Type	Transmission
da Vinci Xi [21,22]	FDA 2000 (orig.)	3 + 3 wristed	Mechanical (parallelogram)	Cable–pulley
Senhance [14]	FDA 2017	3 + 3 wristed	Mechanical (parallelogram)	Cable/gear
Revo-i MSR-5000 [26]	KFDA 2017	3 + 3 wristed	Mechanical (parallelogram)	Cable–pulley
Hugo RAS [15]	CE 2021	6 per arm cart	Software (4 cart arms)	Direct geared
Versius [16,17]	CE 2019; FDA De Novo 2024	7 (modular cart)	Software (7-DoF cobot)	Direct geared
Avatera [30,31]	CE 2019	7 anthropomorphic	Software (7-DoF arms)	Direct geared
Hinotori [4]	PMDA Japan 2020	8 per operating arm	Software (8-axis arms)	Direct geared
Dexter [40,41]	CE 2022	Wrist + cart	Hybrid (spherical wrist + docking)	Concentric tubes + cable jaws
MiroSurge (research) [32]	Research only	7-DoF MIRO arms	Software (torque-controlled)	Cable-driven distal wrist
Raven-II (research) [36]	Research only	7-axis cable spherical	Mechanical (spherical 5-bar)	Cable-driven
Proposed (this work)	Research demonstrator	2 + 1 q_3 (orientation)	Mechanical (spherical)	Contra-rotating differential gearbox

3. Materials and Methods

3.1. Kinematic Model

This section presents the kinematic model that governs the motion of the RCM spherical robotic mechanism. Given the two different peripheral control devices integrated into the system for teleoperating the robot—a 7-DoF haptic device and a 3-axis joystick—two sets of forward and inverse kinematic parametrizations were derived: the first for Cartesian-space control, used when the robot is teleoperated through the haptic device, and the second for simple orientation around the RCM point when using the 3-axis joystick.

The kinematic scheme presented in Figure 1 illustrates the spherical mechanism with six links connected through seven revolute joints, two active and five passive. The virtual axis of each revolute joint is oriented so that all axes meet in the center of the spherical mechanism, establishing a remote center of motion around which the structure rotates. The parameters q_1 , q_2 and q_3 represent the active joints.




Figure 1. Kinematic scheme of the spherical mechanism.

When both links L_1 and L_2 rotate synchronously in opposite directions, the spherical mechanism extends or retracts. As the angle between links L_1 and L_2 shrinks, the same motion is transmitted to links L_5 and L_6 through L_3 and L_4 , and vice versa. When both links L_1 and L_2 rotate synchronously in the same direction with the same angle, the entire mechanism rotates around the base X_B axis. More specifically, if the angle between links L_1 and L_2 remains constant, the entire mechanism can rotate around the base axis X_B . Based on this, a more simplified method of control was devised, where the two actuators are decoupled: one specifically controls the extension and retraction of the arm, while the other controls the rotation of the entire arm around the X_B base axis, essentially eliminating any form of dependency between q_1 and q_2 .

To simplify the process of computing the kinematic equations, the entire structure can be projected from three-dimensional space onto a two-dimensional plane by extending the RCM point to infinity and “unwrapping” the spherical mechanism, thereby obtaining a more intuitive topology that resembles a parallelogram mechanism with scissor mounting, or a jack mechanism, as illustrated by the kinematic scheme in Figure 2.

This is useful in determining the relationships between the angles of each link. From here, the mechanism can be further enhanced by virtually extending the L_1 and L_2 links until they meet links L_5 and L_6 at two different points, marked by joints R_8 and R_9 , as illustrated in Figure 3. This configuration, along with many others, can function properly as long as the links themselves remain parallel to each other, maintaining the form of a parallelogram mechanism.




Figure 2. Kinematic scheme of the spherical mechanism (plane OYZ).




Figure 3. Kinematic scheme of the mechanism with virtual extended links, illustrating the parallelogram-like topology of the spherical structure.

For kinematic modeling, the standard 3×3 rotation matrices around the X , Y and Z axes by angle γ are denoted $R_x(\gamma)$, $R_y(\gamma)$, and $R_z(\gamma)$. Each rotation matrix belongs to the special orthogonal group $SO(3)$, which represents all possible rotations in three-dimensional space around an origin, expressed in 3×3 orthogonal matrices with a determinant of +1, preserving orientation.

3.1.1. Symmetric Modeling of the Spherical Chain

Given the symmetrical nature of the mechanism's architecture, it is sufficient to study only the K_1 kinematic chain for kinematic modeling, which represents one lateral half of the entire mechanism. The mapping between the reference frame and the end-effector frame is described through the following transformation:

$$T_1 = R_X(q_2) \cdot R_X(q_1) \cdot R_Z(\alpha_1) \cdot R_X(u_1) \cdot R_Z(\alpha_1) \quad (1)$$

where $\alpha_1 = 15^\circ$ is the geometric angle of separation between adjacent joint axes with respect to the RCM, q_1 and q_2 are the active joint coordinates and u_1 is the passive joint coordinate of the studied chain. The OX local axis of the frame attached to joint R_1 describes the unit vector associated with the axis of the surgical instrument:

$$OX_1 = T_1 \cdot \begin{bmatrix} 1 \\ 0 \\ 0 \end{bmatrix}, \quad \|OX_1\| = 1 \quad (2)$$

$$(R-L) \cdot OX_1 = \begin{bmatrix} X_E \\ Y_E \\ Z_E \end{bmatrix} \quad (3)$$

$$FK: \begin{cases} X_E = f_1(q_1, L) \\ Y_E = f_2(q_1, q_2, L) \\ Z_E = f_3(q_1, q_2, L) \end{cases} \quad (4)$$

$$IK: \begin{cases} q_1 = g_1(X_E, Y_E, Z_E) \\ q_2 = g_2(X_E, Y_E, Z_E) \\ L = g_3(X_E, Y_E, Z_E) \end{cases} \quad (5)$$

3.1.2. RCM-Orientation Parametrization for Joystick Teleoperation

For the 3-axis joystick, a simpler RCM-orientation parametrization in terms of three orientation angles (ψ, θ, φ) describes the rotation of the instrument axis around the RCM:

$$RCM = [\psi, \theta, \varphi]^T \quad (6)$$

$$RCM_0 = \begin{bmatrix} \psi \\ \theta \\ \varphi \end{bmatrix} = R_X(\psi) \cdot R_Y(\theta) \cdot R_Z(\varphi) \cdot \begin{bmatrix} 1 \\ 0 \\ 0 \end{bmatrix} = OX_1 \quad (7)$$

When the mechanism is used purely for orientation, the insertion length L is decoupled from the other motions, and the roll angle φ is driven exclusively by q_3 , leading to the compact forward and inverse forms:

$$\text{FK}_{\text{RCM}} : \begin{cases} \varphi = q_3 \\ \psi = -\arcsin\left(\frac{2\sin(q_2)\sin^2(q_1)\cos^2(\alpha_1) - 2\cos(q_2)\cos(q_1)\sin(q_1)\cos^2(\alpha_1) - \sin(q_2)}{\sqrt{1 - \sin^2(q_1)\sin^2(2\alpha_1)}}\right) \\ \theta = \text{atan2}\left(2\sqrt{1 - \sin^2(q_1)\sin^2(2\alpha_1)}\sin(2\alpha_1)\sin(q_1), \cos^2(2\alpha_1) + \sin^2(2\alpha_1)\cos(2q_1)\right) \end{cases} \quad (8)$$

$$\text{IK}_{\text{RCM}} : \begin{cases} q_3 = \varphi \\ q_1 = \text{atan2}\left(\text{csgn}\left(\sin\frac{\theta}{2}\right)\sin\frac{\theta}{2}\csc(2\alpha_1), \sqrt{2\cos\theta - 2\cos(4\alpha_1)}\right) \\ q_2 = g_2(\psi, \theta) \end{cases} \quad (9)$$

with associated velocity Jacobians J_3 and J_4 .

3.1.3. Mobility, Jacobian and Singularity Analysis

The mobility of the spherical linkage is established before the differential analysis. The mechanism comprises six moving links interconnected by seven revolute joints whose axes all intersect at the remote center of motion, so that every admissible motion belongs to the rotation group $\text{SO}(3)$. For the extension–retraction loop, which has the topology of a double spherical parallelogram, the spherical form of the Chebyshev–Grübler–Kutzbach criterion yields a single internal degree of freedom, driven by the active joint q_2 , as shown in Equation (10). The base-axis rotation q_1 and the end-effector roll q_3 contribute two further actuated degrees of freedom. The instrument is therefore positioned with three actuated degrees of freedom: two degrees of freedom orient the instrument axis about the remote center of motion, described by the angles ψ and θ , while the third degree of freedom defines the roll angle φ , with $\varphi \equiv q_3$.

$$M = 3(\ell - 1) - 2j = 3(3) - 2(4) = 1 \Rightarrow \underset{q_2}{1} + \underset{q_1}{1} + \underset{q_3}{1} = 3 \text{ actuated DoF} \quad (10)$$

Substituting the rotation operators of Equation (1) into the orientation parametrization of Equations (6) and (7) and eliminating the passive coordinate u_1 through the loop-closure constraint, yields the explicit closed-form forward map of Equation (8). In this map, $\varphi = q_3$ holds exactly, the elevation angle θ depends only on the base-rotation coordinate q_1 , and the azimuth ψ depends on both active coordinates q_1 and q_2 . The corresponding inverse map is given in Equation (9); because $\theta(q_1)$ is strictly monotonic over the operating range, with $\partial\theta/\partial q_1 > 0$ for $|q_1| < 90^\circ$, q_1 is recovered uniquely from θ , after which q_2 follows from ψ in closed form, as shown in Supplementary Materials Appendix A. These explicit expressions replace the general placeholder forms $f_1 \dots f_3$ and $g_1 \dots g_3$ used in the definition of the model and make the decoupling structure precise: the roll axis is exactly decoupled, the elevation θ is governed by a single joint, whereas the azimuth ψ remains kinematically coupled to both active joints.

Differentiating the forward map with respect to the actuated coordinates gives the velocity Jacobian J_3 , which relates the joint rates to the rates of the instrument-orientation parameters, namely $(\dot{q}_1, \dot{q}_2, \dot{q}_3) \mapsto (\dot{\psi}, \dot{\theta}, \dot{\varphi})$. Because $\varphi = q_3$ and θ is independent of q_2 , J_3 is lower-block-triangular, with the closed-form entries given in Equation (11). In particular, $\partial\theta/\partial q_1$ admits the compact form shown in Equation (11), while the azimuth row couples both active joints through $\partial\psi/\partial q_1$ and $\partial\psi/\partial q_2$.

$$J_3 = \frac{\partial(\psi, \theta, \varphi)}{\partial(q_1, q_2, q_3)} = \begin{bmatrix} \frac{\partial\psi}{\partial q_1} & \frac{\partial\psi}{\partial q_2} & 0 \\ \frac{\partial\theta}{\partial q_1} & 0 & 0 \\ 0 & 0 & 1 \end{bmatrix}, \quad \frac{\partial\theta}{\partial q_1} = \frac{2 \sin(2\alpha_1) \cos(q_1)}{\sqrt{1 - \sin^2(q_1) \sin^2(2\alpha_1)}} \quad (11)$$

The determinant of the orientation Jacobian factorizes as $\det J_3 = -(\partial\psi/\partial q_2)(\partial\theta/\partial q_1)$, so the mechanism is singular only when one of these two scalar factors vanishes, as shown in Equation (12). The first factor gives $\partial\theta/\partial q_1 = 0 \Leftrightarrow \cos q_1 = 0$, i.e., $q_1 = \pm 90^\circ$, which corresponds to a boundary singularity at the extreme base-link configuration. The second factor gives $\partial\psi/\partial q_2 = 0 \Leftrightarrow \cos q_2 = 2 \sin q_1 \sin(q_1 + q_2) \cos^2 \alpha_1$, which defines the extension-retraction singular locus. With the design value $\alpha_1 = 15^\circ$, both singularity conditions occur only near the boundary of the reachable region; therefore, the central operating workspace remains singularity-free.

$$\det J_3 = -\frac{\partial\psi}{\partial q_2} \frac{\partial\theta}{\partial q_1} = 0 \Leftrightarrow \underbrace{\cos(q_1) = 0}_{q_1 = \pm 90^\circ} \quad \text{or} \quad \underbrace{\cos(q_2) = 2 \sin(q_1) \sin(q_1 + q_2) \cos^2(\alpha_1)}_{\text{extension/retraction singular locus}} \quad (12)$$

Figure 4 maps the absolute value of the determinant of the orientation Jacobian J_3 over the actuated joint space (Figure 4a) and the corresponding reachable instrument-axis orientations ψ and θ (Figure 4b), both evaluated for $\alpha_1 = 15^\circ$. The red curve in Figure 4a marks the extension and retraction singular locus. Away from this locus and from the boundary configurations, the manipulability index $|\det J_3|$ remains above zero across the selected operating range, confirming smooth and well-conditioned orientation control in the central workspace. Representative forward-kinematics values illustrate this behavior: $(q_1, q_2) = (20^\circ, 10^\circ)$ gives $(\psi, \theta) = (47.50^\circ, 19.69^\circ)$ with $|\det J_3| = 0.954$; $(30^\circ, -15^\circ)$ gives $(\psi, \theta) = (41.57^\circ, 28.96^\circ)$ with $|\det J_3| = 0.894$; and $(-25^\circ, 20^\circ)$ gives $(\psi, \theta) = (-26.99^\circ, -24.40^\circ)$ with $|\det J_3| = 0.927$. These configurations are all located away from singular configurations.




Figure 4. Conditioning and reachable orientation workspace of the spherical remote center-of-motion mechanism for $\alpha_1 = 15^\circ$: (a) absolute determinant of the orientation Jacobian J_3 over the actuated

joint space, with the red curve indicating the extension and retraction singular locus; (b) reachable instrument-axis orientations ψ and θ , colored by the manipulability index $|\det J_3|$.

3.2. Robot Structure Design and CAD Modeling

The complete robotic structure was designed in Siemens NX 2506 (Plano, TX, USA), with the objective of integrating the spherical RCM mechanism, the contra-rotating gearbox, the actuator housings, the supporting bearings and the base frame into a compact proof-of-concept assembly. The design was developed around the requirement that the extension and retraction motion of the spherical mechanism and the rotation around the base axis should be driven by independent actuator chains. Figure 5 presents the overall mechanical structure in two representative configurations: folded (a) and unfolded (b).

Figure 6 illustrates an exploded view of the components integrated into the contra-rotating gearbox, while Figure 7 presents a sectioned view showing how these components mesh together, with each component referenced in Table 3.




Figure 5. Robotic mechanical structure in two representative configurations: (a) folded configuration; (b) unfolded configuration.

The key design challenge was the actuation of the two base links, L_1 and L_2 , during extension and retraction. This motion requires the two links to rotate synchronously in opposite directions. To achieve this mechanically, a contra-rotating gearbox was developed and integrated at the base of the spherical mechanism. The gearbox converts one input rotation into two coaxial output rotations of equal magnitude and opposite direction, allowing the spherical mechanism to extend and retract without requiring cable transmissions or two electronically synchronized motors.

The main components of the contra-rotating gearbox are illustrated in Figure 6a,b. The input shaft also acts as the first output shaft and drives the input bevel gear. This gear meshes with two satellite bevel gears mounted symmetrically inside the gearbox housing. The satellite gears transmit motion to the output bevel gear, which drives the second coaxial output shaft in the opposite direction. In this configuration, the gearbox has a 1:1 transmission ratio, which was sufficient for proof-of-concept validation, although it limits torque amplification and motion resolution.




Figure 6. Contra-rotating gearbox components: (a) exploded view of the input-side components; (b) exploded view of the output-side components.

The gearbox housing was designed to accommodate the input shaft, the second output shaft, the bevel gear pair, the satellite gears and the supporting bearings. The main shafts are supported by 6203-2RS radial ball bearings, while the satellite bevel gear shafts are supported by 6000-2RS radial ball bearings. The first output shaft includes a keyway interface for coupling with link L_1 , while the second output shaft is hollow, allowing the first shaft to pass through it concentrically. To reduce friction between the two coaxial shafts, K10×13×10 radial needle roller cages were placed inside the second output shaft.

Figure 7 illustrates the sectioned view of the contra-rotating gearbox and clarifies the internal meshing between the input bevel gear, the satellite bevel gears and the output bevel gear.




Figure 7. Contra-rotating gearbox section view.

The rotation of the entire spherical mechanism around the base axis is driven separately from the extension and retraction motion. For this purpose, the contra-rotating gearbox assembly is mounted in the base frame and rotated by a second actuator through a helical gear transmission. The q_1 actuator drives a helical gear shaft located below the contra-rotating gearbox, while the corresponding driven helical gears are integrated into the gearbox housing. As a result, the entire gearbox assembly can rotate around the base axis without affecting the opposed rotation of links L_1 and L_2 .

The structural elements associated with this base-axis rotation are presented in Figure 8, including the q_1 actuator, q_2 actuator, actuator housing, helical gear pair, elastic coupler, bearing blocks and robot base frame. The inner bearing block also supports the inductive sensors used for homing and limiting the q_1 motion. The nearby helical gear includes fourteen radial holes spaced at 25° intervals, allowing screws to be positioned as adjustable mechanical references for the sensor-based angular limits.




Figure 8. Elements of the robot structure, showing the contra-rotating gearbox, the helical gear pair driving the q_1 rotation, the base frame and the placement of the two main actuators.

The spherical RCM mechanism consists of links L_1 to L_6 connected through active and passive revolute joints. The axes of the joints are arranged with a 15° angular separation with respect to the RCM point, according to the kinematic architecture described previously. The q_2 homing and end-stop function is implemented through an inductive sensor mounted on link L_2 , which detects a metallic screw placed on link L_3 when the mechanism approaches its retracted limit.

Figure 9 presents a sectioned view of the active and passive joints of the spherical mechanism. The passive revolute joints are assembled around custom aluminum shafts. Axial play is reduced using AXK0619 needle roller cages and AS0619 needle roller washers (SKF, Gothenburg, Sweden), while radial stability is provided by 623-ZZ radial ball bearings (NSK Ltd., Tokyo, Japan). The active end-effector joint R_7 , corresponding to q_3 , integrates the q_3 actuator directly inside the link housing. The actuator drives the end-effector active joint shaft through a rigid coupler, while two 624-2RS radial ball bearings stabilize the shaft. The outer end of this shaft is designed to allow the attachment of an endoscopic camera or another slender instrument.




Figure 9. Sectioned view of the spherical mechanism's active and passive joints. The figure shows the passive joint shaft, radial bearings, axial needle roller elements, end-effector active joint shaft, q_3 actuator, rigid coupler and instrument attachment interface.

After the CAD design was completed, the structural components were exported as STL files for additive manufacturing. The robot links, base frame, motor housings and gearbox housings were manufactured using a Stratasys Fortus 380MC system, while the bevel gears were produced using a Stratasys J5 MediJet printer (Eden Prairie, MN, USA) to obtain improved resolution for the gear teeth. The shafts used in the gearbox and rev-

olute joints were manufactured separately based on technical drawings. The final assembly, illustrated in Figure 10, combined the printed structural components, machined shafts, bearings, actuators and sensors into the complete proof-of-concept robotic mechanism.

Table 3. Main components of the robotic mechanism and their correspondence with the numbered elements in Figures 6 to 9.

Part Nr.	Part Name	Quantity	Part Nr.	Part Name	Quantity
1	Input shaft (first output shaft)	1	23	q1 DC motor (FIT0185, DFRobot, Shanghai, China)	1
2	Input bevel gear	1	24	Elastic coupler for q1 DC motor shaft	1
3	Input shaft keyway key	1	25	Right helical gear	1
3 *	Input shaft keyway slot	1	26	Left helical gear	1
4	6203 2RS radial ball-bearing	3	27	Helical gear shaft	1
5	Elastic coupler for q2 DC motor shaft	2	28	Robot base frame	1
6	q2 DC motor (FIT0185, DFRobot, Shanghai, China)	1	29	M3 self-locking nut	5
7	q2 DC motor housing	1	30	Passive joint shaft	1
8	q2 DC motor housing support shaft	1	31	623 ZZ radial ball bearing	5
9	6000 2RS radial ball-bearing	5	32	AS0619 needle roller washers	12
10	Second output shaft	1	33	AXK0619 needle roller cage	6
11	K10X13X10 radial needle roller cages	2	34	L4 link	1
12	Output bevel gear	1	35	L3 link	1
13	Distancing ring	2	36	End-effector active joint shaft	1
14	Satellite bevel gear	2	37	624 2RS radial ball bearing	2
15	Satellite bevel gear shaft	2	38	M4 self-locking nut	5
16	Contra-rotating gearbox housing	1	39	L5 link	1
17	L2 link	1	40	L6 link	1
18	AXK1730 needle roller cage	1	41	Rigid coupler for q3 DC motor shaft	1
19	AS1730 needle roller washer	1	42	q3 DC motor (298:1 Micro Metal Gearmotor HPCB 12 V, Pololu Corporation, Las Vegas, NV, USA)	1
20	L1 link	1	43	q3 DC motor housing	1
21	M8 self-locking nut	4	44 *	Contra-rotating gearbox assembly	1
22	Detachable bearing block (q1 actuator support)	1	45 *	Spherical mechanism assembly	1




Figure 10. Final prototype and gearbox detail: (a) assembled robotic prototype; (b) detail of the contra-rotating gearbox.

3.3. Embedded Hardware and Control Architecture

The mechanical prototype was integrated with a custom embedded hardware and control architecture designed to actuate the three motorized axes, acquire position and limit feedback, communicate with the master console and support the teleoperation workflow used during the phantom-torso experiments. Since the objective of the present work is to report the complete technological demonstrator rather than the detailed electrical design of each circuit block, the electronic subsystem is presented at a functional level. The complete schematic sheets of the microcontroller core, motor drivers, sensor interface and USB–USART communication circuit can be provided as Supplementary Material.

Figure 11 presents the system-level electrical and control architecture of the robotic demonstrator. The host computer runs the graphical user interface, processes the commands received from the selected input device and sends the corresponding joint references to the embedded controller through a USB–USART communication link. Two master devices were integrated: the Force Dimension omega.7 haptic device (Force Dimension, Nyon, Switzerland), used for Cartesian-style teleoperation, and a 3-axis joystick (Guillemot Corporation, Carentoir, France), used for direct RCM-orientation control. The same graphical interface also displays the video stream acquired from the USB endoscopic camera mounted at the end-effector.




Figure 11. System-level electrical and control architecture of the robot.

The custom controller was implemented as a four-layer printed circuit board based on an ATmega2560 microcontroller (Microchip Technology Inc., Chandler, AZ, USA). The board integrates the main subsystems required by the prototype, including motor-driver interfaces, encoder inputs, inductive end-stop inputs, analog current acquisition and USB-USART communication with the host computer. The embedded controller executes the low-level motor-control tasks locally, while the master console remains responsible for teleoperation, inverse kinematics, user interaction and visual feedback.

The main axes of the robot, q_1 and q_2 , are driven through H-bridge motor-driver stages based on IR2104 (Infineon Technologies AG, Neubiberg, Germany) high- and low-side gate drivers and IRF540 N-channel MOSFETs (Infineon Technologies AG, Neubiberg, Germany). These stages allow bidirectional control of the brushed DC gearmotors through PWM and direction commands generated by the microcontroller. The third axis, q_3 , which actuates the end-effector rotation, is driven by a Toshiba TB67H420FTG (Toshiba Electronic Devices & Storage Corporation, Tokyo, Japan) [43] auxiliary motor driver. Position feedback is provided by Hall-effect quadrature encoders, while homing and end-stop detection are implemented using Lanbao inductive proximity sensors (Shanghai Lanbao Sensing Technology Co., Ltd., Shanghai, China). Current feedback is acquired using ACS712 (Allegro MicroSystems, LLC, Manchester, NH, USA) [43] Hall-effect current sensors, allowing abnormal motor-current conditions to be detected during operation.

The assembled controller board is shown in Figure 12. The board provided the hardware interface between the master console and the robotic mechanism, integrating the microcontroller, motor-driver electronics, sensor-conditioning elements and communication interface into a single compact platform. During the experiments, the controller enabled closed-loop position control of the three actuated axes, homing using the inductive sensors and basic protection through current monitoring.




Figure 12. The assembled custom embedded controller board used in the robot.

3.4. Embedded Firmware and Master-Console Interface

The embedded firmware and the master-console software were developed to connect the mechanical prototype, the custom controller board and the operator interface into a functional teleoperated robot, as illustrated in Figure 13. The firmware was implemented in C23 using a bare-metal programming approach on the ATmega2560 microcontroller (Microchip Technology Inc., Chandler, AZ, USA). This approach was selected to allow direct control of the microcontroller peripherals required for real-time operation, including PWM generation, USART communication, ADC acquisition, timer configuration and external interrupt handling.




Figure 13. Embedded firmware architecture and closed-loop motor-control workflow implemented on the ATmega2560 controller (Microchip Technology Inc., Chandler, AZ, USA).

The firmware was organized into modular software blocks corresponding to the main embedded functions of the system. The main module coordinates the control loop and processes the commands received from the host computer. The configuration module centralizes the hardware definitions and timing parameters. The USART module manages serial communication with the master console. The sensor module handles ADC acquisition for the current sensors and configures the external interrupts used for encoder and end-stop feedback. The motor-control module generates the direction and PWM commands required to drive the actuators. This modular organization simplifies debugging and separates communication, sensing and motor-control tasks.

During a surgical procedure, the graphical interface sends target-position commands to the embedded controller through the USB-USART link. The controller reads the encoder counters, verifies the inductive end-stop states, samples the ACS712 [44] current sensors and computes the motor commands using a position-based PID controller. The control output determines both the rotation direction and the PWM duty cycle applied to each motor driver. Current monitoring is used as a basic protection layer: if abnormal motor current is detected, the corresponding motor command can be stopped to prevent continued motion against an obstacle or mechanical limit.

To address the motion resolution constraints imposed by the 1:1 differential gearbox ratio and the inherent non-linearities of the 3D-printed gearing, an AI-based control upgrade path is envisioned for the platform. Instead of relying solely on static PID coefficients, an Adaptive Neuro-Fuzzy Inference System (ANFIS) or a Reinforcement Learning (RL) auto-tuning framework can be integrated. By utilizing the real-time motor current feedback from the ACS712 sensors alongside encoder telemetry, such an intelligent controller can dynamically adapt the PWM duty cycles to counteract varying frictional resistances and torque drops in mechanically unfavorable orientations, ensuring smoother trajectory tracking around the RCM.

The master-console application was developed in C++/CLI (Microsoft Visual Studio 2022, version 17.12, Microsoft Corporation, Redmond, WA, USA) using Windows Forms .NET 10.0. Its role is to provide the operator with a single interface for selecting the communication port, connecting the robot, selecting the active peripheral control device, starting the homing routine, engaging teleoperation and displaying the endoscopic video stream. The GUI also performs higher-level motion processing: it acquires data from the selected master device, applies the corresponding kinematic mapping and transmits the resulting motor references to the embedded controller.


Two teleoperation modes were implemented. In the haptic-device mode, the Force Dimension omega.7 (Force Dimension, Nyon, Switzerland) device is used to generate Cartesian-style commands that are converted into joint references through the inverse kinematic model. In the joystick mode, the 3-axis joystick is used for RCM-orientation control, where the joystick axes directly command the orientation of the endoscopic camera around the remote center of motion. The joystick-based mode is particularly suitable for camera-orientation tasks and was used during the phantom-torso validation.

The video feedback subsystem was implemented using OpenCV 4.13.0. The USB endoscopic camera mounted at the end-effector provides the image stream, which is displayed in the graphical interface during teleoperation.

Furthermore, the OpenCV-based visual pipeline serves as the foundation for an intelligent computer vision subsystem. By deploying a lightweight deep learning object detection network on the host computer, the system can automatically segment and track the internal visual markers in real time. This AI-driven feature extraction transforms the camera from a passive feedback tool into an active component of a closed-loop visual servoing architecture, capable of calculating alignment errors and generating autonomous



corrective commands to keep targeted features centered without continuous manual joystick adjustments.

The user interface was designed to be easy to comprehend, having two main tabs, namely the connection and control tabs, where important functionalities are located. Figure 14a below illustrates these controls as follows: (1) Main tab selection menu; (2) Button menu for robot teleoperation start/stop, robot homing, camera feedback start/stop; (3) Motion setting group, with three checkboxes corresponding to each axis of the joystick/haptic device and a numeric field to set the motion increment; and (4) Incremental control group, comprising four buttons for the X and Y axis of teleoperation. Figure 14b illustrates the marker pad.



(a)

Markers PAD



(b)

Figure 14. Master-console graphical user interface and teleoperation workflow: (a) control interface for motion settings and teleoperation; (b) endoscopic video feedback displayed through the OpenCV stream.

3.5. Phantom Torso Experimental Setup

The surgical robot was evaluated through a qualitative phantom-torso experiment intended to verify the functional feasibility of the spherical RCM mechanism for endoscopic camera-orientation tasks. The objective of this experiment was to determine whether the assembled system could be teleoperated around the remote center-of-motion point while providing useful visual feedback from inside a constrained, torso-like workspace.

The experimental setup presented in Figure 15 consisted of a phantom torso, a 3D-printed custom pad with an array of markers, a custom 3D-printed endoscopic camera, and the robot itself. The phantom torso was used to provide an anatomical reference similar to a minimally invasive surgical environment, constraining the range of motion of the mechanism to a limited workspace. The marker pad was modeled to have slightly concave geometry to allow the endoscopic camera to roughly center each marker visually on the graphical user interface video stream.




Figure 15. Experimental setup—robotic system and phantom torso.

The internal marker pad was placed inside the phantom torso in a position where the camera could visualize the individual markers during the orientation of the spherical mechanism. The pad was designed with slightly concave geometry to facilitate visual centering of the markers in the camera image. During the experiment, the operator attempted to orient the endoscopic camera so that each marker appeared successively in the graphical interface video stream. This setup made it possible to test the practical interaction between the RCM mechanism, the teleoperation interface, the embedded controller and the video-feedback subsystem.

Before the experiment, the robot was powered on and connected to the master console through the serial interface. The automatic homing routine was then executed, using the inductive end-stop sensors to define the reference positions of the actuated axes. After homing, the endoscopic video stream was started in the graphical user interface. The joystick-based RCM-orientation mode was used as the main teleoperation method because it provided a direct mapping between operator input and camera orientation around the remote center of motion. The Force Dimension omega.7 haptic device remained available as an alternative input device for Cartesian-style control, but the phantom-torso evaluation focused primarily on camera orientation using the joystick.

4. Results

The robot was manufactured, assembled and integrated into a functional teleoperated platform. The final system included the spherical RCM mechanism, the contra-rotating gearbox, the helical-gear transmission for base-axis rotation, the end-effector rotation axis, the inductive end-stop sensors, the custom four-layer controller board, the embedded firmware, the master-console graphical interface and the custom USB endoscopic camera.

The mechanical structure was able to reach both extended and retracted configurations during laboratory testing. The contra-rotating gearbox enabled the opposed rotation of the two base links during extension and retraction, while the second actuator chain rotated the gearbox assembly around the base axis. The end-effector actuator provided additional camera-orientation capability. The custom controller board enabled motor actuation, encoder-based feedback, inductive-sensor homing and current monitoring during operation.

The phantom-torso experiment confirmed that the system could be teleoperated. After the homing routine, the operator used the joystick-based RCM-orientation mode to move the endoscopic camera around the torso entry point. The camera was oriented from left to right and from a more extended configuration toward a more retracted configuration, while the internal marker pad was observed through the graphical interface.

The video feedback from the USB endoscopic camera was displayed successfully in the master-console interface. The operator was able to bring individual markers into the camera's field of view and visually center them on the screen. This confirmed that the mechanical structure, embedded controller, joystick teleoperation and video-feedback subsystem could operate together during a representative laboratory task.

The experiment demonstrated that the proposed spherical RCM architecture can be used as a platform for endoscopic camera-orientation tasks. The robot responded to joystick commands and allowed the camera to be repositioned around the phantom-torso access point. The marker pad could be observed from several camera orientations, indicating that the proposed architecture provides a usable range of motion for qualitative visual targeting.

However, the evaluation also revealed several technical limitations. The first limitation was related to motion resolution. Because of the 1:1 ratio of the contra-rotating gearbox and the characteristics of DC motor transmission, small motor motions could produce relatively large visible changes in camera orientation. This made precise marker centering more difficult in some configurations.

The second limitation was related to the available actuation torque. In certain orientations, especially when the mechanism was positioned close to the phantom torso or when the configuration became mechanically less favorable, the motors had reduced ability to move the structure smoothly. This limitation is partly associated with the proof-of-concept nature of the 3D-printed gearbox and with the absence of a reduction stage for torque amplification.

The third limitation was observed in retracted configurations of the spherical mechanism. When the arm was retracted, the links formed a more compact semi-spherical shape, reducing the usable rotation around the base axis because of possible interference with the phantom torso. In contrast, when the mechanism was more extended, the available angular motion around the base axis increased.

The results of the tests are depicted in Table 4.

Table 4. Qualitative and quantitative experimental observations and identified limitations during the phantom torso evaluation.

Aspect Evaluated	Qualitative Observation	Limitation/Proposed Mitigation
Synchronized opposed rotation of L_1 , L_2	Achieved consistently from the single q_2 input	None observed (mechanical synchronization)
Decoupling q_1 and q_2	No visible coupling during pure q_1 or pure q_2 motions	None observed
Marker centering on GUI video stream	Each marker brought into the frame; fine centering sometimes difficult	High motion resolution from 1:1 gearbox + 3D-printed gearing, use higher-ratio gearbox + metal-cut gears
Torque on phantom torso	Occasionally insufficient to overcome foam resistance in adverse orientations	Stronger motor or higher-ratio gearbox
Workspace in retracted configuration	Reduced angular range of q_1 due to torso proximity	Reposition the robot above the patient or at a steeper lateral angle
Visual stability of the trocar point	OptiTrack evaluation confirmed that the remote center of motion was maintained with 1.69 mm mean deviation and 2.13 mm RMS deviation during the analyzed orientation sweep	Further tests should include repeated trials, formal repeatability evaluation and complete workspace characterization
Teleoperation responsiveness	Both control modes (joystick/haptic) responded well to operator input	Latency was not quantitatively measured; future work to characterize end-to-end loop time

To complement the qualitative phantom-torso evaluation, the remote-center-of-motion accuracy was measured using an OptiTrack optical tracking system, as illustrated in Figure 16.




Figure 16. OptiTrack (NaturalPoint, Inc., Corvallis, OR, USA) measurement setup.

Five retro-reflective markers were mounted collinearly on the external shaft of the instrument, as illustrated in Figure 17, on the same rigid segment that transmits motion to the part passing through the trocar. The robot was commanded through an orientation sweep around the remote center of motion, while the three-dimensional marker trajectories were recorded.




Figure 17. Instrument with the five collinear retro-reflective markers mounted on the shaft that holds the remote center of motion.

Because the five markers were mounted in series along the instrument shaft, the rigid-body orientation returned by the motion-capture software was not used directly. A collinear marker arrangement is poorly conditioned for estimating rotation about the shaft axis. Instead, the remote center of motion was reconstructed directly from the marker po-

sitions. During an ideal pivoting motion, each marker fixed on the shaft describes a spherical trajectory whose center coincides with the remote center of motion. Therefore, the remote center was estimated as the common center of the marker trajectories using a least-squares sphere-center fitting procedure.

The five markers mounted on the instrument shaft were used for the analysis. Other markers detected by the OptiTrack system were excluded, since they did not belong to the instrument marker set and produced inconsistent or intermittent trajectories. Frames affected by optical-tracking dropouts, marker swaps or residual outliers were removed before computing the final accuracy metrics. After this filtering step, 9915 valid frames were retained for the final RCM accuracy analysis. The outlier rejection applied a fixed per-frame threshold of approximately three times the RMS deviation (6.5 mm), corresponding to a 3σ criterion.

The reconstructed remote center of motion was located at approximately $[-466, 200, -41]$ mm in the OptiTrack global coordinate frame. The measured RCM positional deviation had a mean value of 1.69 mm, a root-mean-square value of 2.13 mm, a median value of 1.51 mm and a 95th percentile of 4.39 mm. These results indicate that the proposed contra-rotating differential mechanism maintained the remote-center-of-motion constraint within approximately 2 mm RMS during the analyzed orientation sweep. The obtained values are close to the expected optical-tracking uncertainty of the marker cluster, suggesting that the reported deviation represents an upper bound on the actual mechanical RCM error of the proof-of-concept prototype.

Because the remote center was reconstructed by least-squares sphere fitting from five collinear markers, the reported RMS deviation combines the true mechanical wander of the pivot with the positional noise of the optical-tracking system and the reconstruction error of the ill-conditioned collinear geometry. It is therefore of the same order as the noise floor of the measurement chain and represents an upper bound on the mechanical RCM error rather than a sub-millimeter precision. Repeatability was not characterized in this proof-of-concept study, which was limited to a single continuous sweep. The planned protocol, executing the same commanded sweep N times under identical homing and reporting the standard deviation of the independently reconstructed center positions across repetitions, will be applied to the redesigned prototype.

Quantitative evaluation of remote-center-of-motion accuracy are presented in Table 5 and Figure 18.

Table 5. OptiTrack-based quantitative evaluation of remote center-of-motion accuracy.

Metric	Symbol	Units	Measured Value
RCM positional deviation, mean	$\bar{e}_{RCM}$	mm	1.69
RCM positional deviation, RMS	e_{RMS}	mm	2.13
RCM positional deviation, median	e_{med}	mm	1.51
RCM positional deviation, 95th percentile	e_{95}	mm	4.39
Valid analyzed frames	N	-	9.915
Per-frame outlier threshold ($\approx 3 \times RMS$)	-	mm	6.5




Figure 18. OptiTrack-based reconstruction of the remote center of motion: (a) representative marker trajectories recorded during the orientation sweep and reconstructed remote center of motion; (b) distribution of the RCM positional deviation after removing optical-tracking dropouts and residual outliers.

5. Discussion

The present work demonstrates the development of a complete proof-of-concept robot based on a mechanically constrained spherical remote center-of-motion mechanism for minimally invasive surgical orientation tasks. Unlike purely conceptual mechanism studies, the proposed system integrates the mechanical structure, the contra-rotating gear-box, the custom embedded controller, the firmware, the graphical interface and the endoscopic video-feedback subsystem into a functional robot. The phantom-torso experiment showed that the mechanism can be teleoperated to orient an endoscopic camera inside a constrained workspace and to visualize internal markers through the graphical interface. The validation combines a qualitative phantom-torso demonstration with a quantitative optical-tracking measurement of RCM accuracy and should be interpreted as an initial feasibility demonstration rather than a complete performance assessment or clinical validation.

5.1. Relevance of the Mechanically Constrained Spherical RCM Architecture

The principal advantage of mechanically constrained RCM mechanisms is that the pivoting behavior is imposed by the geometry of the mechanism itself. In the context of minimally invasive surgery, this is important because the instrument or camera shaft must rotate around the trocar region without generating excessive lateral displacement at the incision. Software-constrained RCM approaches can provide greater flexibility because the pivot point can be adapted through control algorithms, but they depend on the accuracy of kinematic calibration, sensor feedback, and real-time controller performance. In contrast, a mechanically constrained architecture offers a deterministic pivoting behavior that is less dependent on continuous software enforcement.

The spherical architecture investigated in this work belongs to this mechanically constrained family. Its revolute joint axes are arranged to intersect at a common remote center, allowing the end-effector orientation to be modified around the RCM point. Compared with planar parallelogram mechanisms, a spherical linkage can provide a compact geometric arrangement around the trocar region. This may be useful in future multi-arm configurations, where several manipulators must operate around the same patient-side workspace. At the same time, spherical mechanisms are more demanding in terms of mechanical design, manufacturing accuracy, and assembly alignment because small deviations in joint-axis positioning can affect the practical quality of the RCM behavior.

5.2. Contra-Rotating Gearbox as the Main Design Contribution

The contra-rotating gearbox represents the main mechanical contribution of the proposed prototype. In the spherical RCM mechanism, L_1 and L_2 must rotate synchronously in opposite directions during extension and retraction. A possible solution would be to use two independent motors and synchronize them through software, but this increases the number of actuators and introduces sensitivity to motor mismatch, controller errors, and calibration drift. Another solution, used in several spherical or tensegrity RCM concepts, is cable or pulley actuation, but this introduces issues related to cable pretension, elongation, friction, and periodic readjustment.

The proposed gearbox solves this synchronization problem mechanically. A single input drives two coaxial outputs with equal magnitude and opposite direction, thereby producing the required opposed rotation of the two base links. This reduces the need for electronic synchronization between two base-link actuators and eliminates cable-based transmission for the extension and retraction motion. In addition, the gearbox allows the extension-retraction motion of the spherical mechanism to be mechanically separated from the rotation of the entire assembly around the base axis. This decoupling simplifies both mechanical interpretation and control because one actuator chain commands the spherical extension–retraction motion while another commands the base-axis rotation.

The current gearbox implementation also introduces important limitations. The prototype uses a 1:1 differential ratio, which means that the angular output of the base links is directly related to the motor output without torque amplification or fine angular reduction. This contributed to reduced fine-positioning capability during marker centering in the phantom experiment. Moreover, the bevel and helical gears were 3D printed, which is appropriate for proof-of-concept validation but not optimal for backlash, stiffness, wear, or torque transmission. A future version should therefore introduce a non-unitary reduction ratio and replace the critical printed gear elements with machined metal components.

5.3. Embedded Control and Teleoperation Integration

A relevant aspect of the work is that the mechanism was not evaluated as a passive mechanical model, but as an active robot. The custom four-layer controller board, the ATmega2560-based firmware, the motor-driver stages, the encoder feedback, the inductive end-stops, the current sensors, and the graphical user interface were all integrated.

The embedded controller allowed local closed-loop position control of the three actuated axes, while the computer handled the higher-level teleoperation interface, kinematic command generation, and video display. This separation is appropriate because it keeps the time-sensitive motor-control tasks on the microcontroller and the user-facing functions on the computer. The use of both a haptic device and a joystick also demonstrates that the mechanism can support different command parametrizations: Cartesian-style control through inverse kinematics and direct RCM-orientation control through the joystick.

For the phantom-torso experiment, joystick-based orientation control was the most practical approach because the experimental task focused on camera orientation rather than full instrument positioning. The graphical interface and OpenCV-based camera feed-back allowed the operator to visualize the internal marker pad and guide the camera orientation accordingly. However, the current embedded architecture remains a proof-of-concept implementation. The 8-bit ATmega2560 is sufficient for basic PID control and serial communication, but future versions would benefit from a 32-bit microcontroller, higher control-loop frequency, improved current sensing, real-time communication protocols, and more advanced safety monitoring.

5.4. Interpretation of the Phantom-Torso Evaluation

The phantom-torso experiment demonstrated that the complete system can be assembled, initialized, teleoperated and used to orient an endoscopic camera toward internal visual targets. The experiment also confirmed that the system components—mechanical structure, gearbox, motors, sensors, controller board, firmware, GUI and camera feed-back—can operate together during a representative laboratory task. In this sense, the validation supports the feasibility of the proposed spherical RCM architecture.

The phantom-torso experiment itself was qualitative, but it was complemented by a quantitative optical-tracking measurement of the remote center of motion. Using the OptiTrack system, the pivot constraint was maintained with a mean positional deviation of 1.69 mm and an RMS deviation of 2.13 mm over the analyzed orientation sweep, values close to the optical-tracking noise floor of the marker cluster, and therefore an upper bound on the true mechanical RCM error. Repeatability across repeated trials, the full angular workspace, marker-centering accuracy and dynamic response were nonetheless not characterized. The results should therefore not be interpreted as proof of surgical accuracy or readiness, but they show that the architecture is mechanically and electronically functional, holds the RCM to within approximately 2 mm RMS, and provides a sound basis for more rigorous quantitative testing.

5.5. Limitations of the Current Prototype

Several limitations must be acknowledged. First, the current prototype does not include a dedicated insertion–retraction axis. The mechanism can orient a camera or slender instrument around the RCM point, but it does not provide full instrument translation along the shaft. Integration of a prismatic insertion axis is necessary before the platform can support more general minimally invasive manipulation tasks.

Second, the actuation system has limited available torque. The current gearbox does not amplify the motor torque because of its 1:1 ratio, and the printed gears further reduce the effective stiffness and transmission quality. A redesigned gearbox with a reduction stage would improve both torque transmission and fine angular control.

Third, the 3D-printed mechanical implementation introduces backlash, surface roughness and dimensional tolerances that are acceptable for laboratory use but insufficient for precise surgical instrumentation. While additive manufacturing is useful for rapid prototyping, future iterations should use machined metallic gears, improved bearing interfaces and higher-stiffness structural elements.

Fourth, the validation was performed only on a foam phantom torso. No biological tissue, ex vivo model, animal model or clinical setting was used. Therefore, no conclusions can be drawn regarding clinical performance, sterilization, biocompatibility, surgical safety or patient-related outcomes. The present contribution is technological and experimental at a proof-of-concept level [45].

5.6. Future Work

Future work will focus on the mechanical improvement and quantitative validation of the system. Priorities include the redesign of the contra-rotating gearbox to include a non-unitary reduction ratio [43–46], replacing the 3D-printed gears with machined metal components [47–52], integrating a dedicated insertion and retraction axis at the end-effector, and upgrading the embedded controller to a 32-bit architecture. Building on the optical-tracking measurement of RCM error reported here, the next prototype will be evaluated across repeated trials to establish positioning repeatability and to map the full angular workspace.

Data-driven compensation of the residual backlash, friction and motion-resolution limits will also be investigated, using ANFIS or reinforcement-learning-based controller auto-tuning trained on the encoder and current data to improve trajectory tracking and RCM accuracy on the existing hardware.

6. Conclusions

This paper presented the development of a compact, architecturally constrained spherical remote center-of-motion robot intended for minimally invasive surgical orientation tasks. The proposed system integrated a spherical RCM mechanism, a contra-rotating gearbox, a custom embedded controller, closed-loop motor control, a master-console interface, and endoscopic video feedback. The main mechanical contribution is the use of a contra-rotating differential gearbox to actuate the opposed motion of the two-base links, eliminating the need for cable-pulley transmissions or software-level synchronization.

The integrated system was evaluated in a phantom-torso setup using a custom endoscopic camera and internal visual markers, confirming the functional integration of all subsystems, and its RCM accuracy was quantified with an OptiTrack optical-tracking system, which showed that the pivot constraint was maintained with a mean deviation of 1.69 mm and an RMS deviation of 2.13 mm. The main remaining limitations are the limited actuator torque and the backlash introduced by the 3D-printed gears. Future work will address these design limitations, add an instrument insertion axis, and extend the optical-tracking evaluation to repeated-trial repeatability and full workspace characterization, as required before advanced surgical evaluations.

Supplementary Materials: The following supporting information can be downloaded at: <https://www.mdpi.com/article/10.3390/technologies14070440/s1>; Video S1: Demonstration; Appendix A.

Author Contributions: Conceptualization, C.V. and P.T.; Methodology, C.V., D.H., I.Z., B.G., A.P., D.C., N.A.H. and D.P.; Software, D.H., I.Z. and P.T.; Validation, C.V., D.H., I.Z., M.M., B.G., C.R., P.T., S.V., D.S., A.P., N.A.H. and D.P.; Formal Analysis, C.V., D.H., M.M., B.G., C.R., S.V., D.S., A.P., D.C., N.A.H. and D.P.; Investigation, D.H., I.Z. and P.T.; Resources, D.P.; Data Curation, M.M., C.R., S.V., D.S. and D.C.; Writing—Original Draft Preparation, C.V., D.H., M.M., B.G., P.T., S.V. and D.S.; Writing—Review and Editing, C.V., I.Z., C.R., D.C. and D.P.; Visualization, I.Z.; Supervision, A.P., D.C., N.A.H. and D.P.; Project Administration, D.P.; Funding Acquisition, D.P. All authors have read and agreed to the published version of the manuscript.

Funding: This work was funded by the project “Romanian Hub for Artificial Intelligence-HRIA”, under the Smart Growth, Digitization and Financial Instruments Program, MySMIS no. 351416.

Informed Consent Statement: Not applicable.

Data Availability Statement: The CAD models, kinematic scripts, AVR firmware source code, GUI source code and experimental videos can be made available from the corresponding author upon reasonable request.

Acknowledgments: This work was funded by the project “Romanian Hub for Artificial Intelligence-HRIA”, under the Smart Growth, Digitization and Financial Instruments Program, MySMIS no. 351416.

Conflicts of Interest: The authors declare no conflicts of interest.

References

1. Bray, F.; Laversanne, M.; Sung, H.; Ferlay, J.; Siegel, R.L.; Soerjomataram, I.; Jemal, A. Global cancer statistics 2022: GLOBOCAN estimates of incidence and mortality worldwide for 36 cancers in 185 countries. *CA Cancer J. Clin.* **2024**, *74*, 229–263. <https://doi.org/10.3322/caac.21834>.
2. Pisla, D.; Birlescu, I.; Vaida, C.; Tucan, P.; Pisla, A.; Gherman, B.; Crisan, N.; Plitea, N. Algebraic modeling of kinematics and singularities for a prostate biopsy parallel robot. *Proc. Rom. Acad. Ser. A* **2019**, *19*, 489–497.
3. Teng, Y.; Xia, C.; Cao, M.; Yang, F.; Yan, X.; He, S.; Cao, M.; Zhang, S.; Li, Q.; Tan, N.; et al. Esophageal cancer global burden profiles, trends, and contributors. *Cancer Biol. Med.* **2024**, *21*, 656–666. <https://doi.org/10.20892/j.issn.2095-3941.2024.0145>.
4. Zhao, Y.X.; Zhao, H.P.; Zhao, M.Y.; Yu, Y.; Qi, X.; Wang, J.H.; Lv, J. Latest insights into the global epidemiological features, screening, early diagnosis and prognosis prediction of esophageal squamous cell carcinoma. *World J. Gastroenterol.* **2024**, *30*, 2638–2656. <https://doi.org/10.3748/wjg.v30.i20.2638>.
5. Morgan, E.; Soerjomataram, I.; Rumgay, H.; Coleman, H.G.; Thrift, A.P.; Vignat, J.; Laversanne, M.; Ferlay, J.; Arnold, M. The Global Landscape of Esophageal Squamous Cell Carcinoma and Esophageal Adenocarcinoma Incidence and Mortality in 2020 and Projections to 2040: New Estimates From GLOBOCAN 2020. *Gastroenterology* **2022**, *163*, 649–658.e2. <https://doi.org/10.1053/j.gastro.2022.05.054>.
6. Stüben, B.-O.; Plitzko, G.A.; Stern, L.; Schmeding, R.; Karstens, K.-F.; Reeh, M.; Treckmann, J.W.; Izbicki, J.R.; Saner, F.H.; Neuhaus, J.P.; et al. Risk Factor Analysis for Developing Major Complications Following Esophageal Surgery—A Two-Center Study. *J. Clin. Med.* **2024**, *13*, 1137. <https://doi.org/10.3390/jcm13041137>.
7. Okamoto, K.; Inaki, N.; Saito, H.; Sakimura, Y.; Tawara, H.; Iino, M.; Tagaya, H.; Doden, K. Analysis of factors associated with operative difficulty in thoracoscopic esophageal cancer surgery in the left-decubitus position: A single-center retrospective study. *BMC Surg.* **2023**, *23*, 242. <https://doi.org/10.1186/s12893-023-02131-2>.
8. Deboever, N.; Jones, C.M.; Yamashita, K.; Ajani, J.A.; Hofstetter, W.L. Advances in diagnosis and management of cancer of the esophagus. *BMJ* **2024**, *385*, e074962. <https://doi.org/10.1136/bmj-2023-074962>.
9. Birlescu, I.; Husty, M.; Vaida, C.; Plitea, N.; Nayak, A.; Pisla, D. Complete Geometric Analysis Using the Study SE(3) Parameters for a Novel, Minimally Invasive Robot Used in Liver Cancer Treatment. *Symmetry* **2019**, *11*, 1491. <https://doi.org/10.3390/sym11121491>.
10. Shemmeri, E.; Wee, J.O. Robotics and minimally invasive esophageal surgery. *Ann. Transl. Med.* **2021**, *9*, 898. <https://doi.org/10.21037/atm-20-4138>.
11. Saeidi, H.; Opfermann, J.D.; Kam, M.; Wei, S.; Leonard, S.; Hsieh, M.H.; Kang, J.U.; Krieger, A. Autonomous robotic laparoscopic surgery for intestinal anastomosis. *Sci. Robot.* **2022**, *7*, eabj2908. <https://doi.org/10.1126/scirobotics.abj2908>.
12. Kim, J.W.; Chen, J.T.; Hansen, P.; Shi, L.X.; Goldenberg, A.; Schmidgall, S.; Scheikl, P.M.; Deguet, A.; White, B.M.; Tsai, D.R.; et al. SRT-H: A hierarchical framework for autonomous surgery via language-conditioned imitation learning. *Sci. Robot.* **2025**, *10*, eadt5254. <https://doi.org/10.1126/scirobotics.adt5254>.
13. Schmidgall, S.; Opfermann, J.D.; Kim, J.W.; Krieger, A. Will your next surgeon be a robot? Autonomy and AI in robotic surgery. *Sci. Robot.* **2025**, *10*, eadt0187. <https://doi.org/10.1126/scirobotics.adt0187>.
14. Stark, M.; Pomati, S.; D’Antonio, A.; Giraudi, F.; Gidaro, S. New telesurgical platform—Preliminary results. *Minim. Invasive Ther. Allied Technol.* **2015**, *24*, 31–36. <https://doi.org/10.3109/13645706.2014.1003945>.
15. Bravi, C.A.; Paciotti, M.; Sarchi, L.; Mottaran, A.; Nocera, L.; Farinha, R.; De Backer, P.; Vinckier, M.-H.; De Naeyer, G.; D’Hondt, F.; et al. Hugo robotic system experience. *Eur. Urol.* **2022**, *82*, 233–237. <https://doi.org/10.1016/j.eururo.2022.04.029>.
16. Thomas, B.C.; Slack, M.; Hussain, M.; Barber, N.; Pradhan, A.; Dinneen, E.; Stewart, G.D. Versius system evaluation. *Eur. Urol. Focus* **2021**, *7*, 444–452. <https://doi.org/10.1016/j.euf.2020.01.011>.
17. FDA. Versius Surgical System (DEN230078). Available online: https://www.accessdata.fda.gov/cdrh_docs/reviews/DEN230078.pdf (accessed on 11 May 2026).

18. Hinata, N.; Yamaguchi, R.; Kusuhara, Y.; Kanayama, H.; Kohjimoto, Y.; Hara, I.; Fujisawa, M. Hinotori Surgical Robot System, a novel robot-assisted surgical platform: Preclinical and clinical evaluation. *Int. J. Urol.* **2022**, *29*, 1213–1220. <https://doi.org/10.1111/iju.14973>.
19. Zhou, P.; Yao, J.; Zhu, S.; Zhang, X.; Zhao, Y. A new remote centre of motion mechanism for robot-assisted minimally invasive surgery. *BioMed. Eng. OnLine* **2018**, *17*, 170. <https://doi.org/10.1186/s12938-018-0601-6>.
20. Pisla, D.; Pusca, A.; Gherman, B.; Pisla, A.; Birlescu, I.; Tucan, P.; Vaida, C.; and Chablat, D. Modeling and Simulation of a Novel Parallel Robotic System for Minimally Invasive Pancreatic Surgery. *ASME. J. Mech. Robot.* **2026**, *18*, 024501. <https://doi.org/10.1115/1.4069975>.
21. Vaida, C.; Birlescu, I.; Gherman, B.; Condurache, D.; Chablat, D.; Pisla, D. An analysis of higher-order kinematics formalisms for an innovative surgical parallel robot. *Mech. Mach. Theory* **2025**, *209*, 105986. <https://doi.org/10.1016/j.mechmachtheory.2025.105986>.
22. Freschi, C.; Ferrari, V.; Melfi, F.; Ferrari, M.; Mosca, F.; Cuschieri, A. Review of the da Vinci system. *Int. J. Med. Robot.* **2013**, *9*, 396–406. <https://doi.org/10.1002/rcs.1468>.
23. Bai, M.; Zhang, M.; Zhang, H.; Pang, L.; Zhao, J.; Gao, C. An Error Compensation Method for Surgical Robot Based on RCM Mechanism. *IEEE Access* **2021**, *9*, 140747–140758. <https://doi.org/10.1109/ACCESS.2021.3117350>.
24. Taylor, R.H.; Funda, J.; Eldridge, B.; Gomory, S.; Gruben, K.; LaRose, D.; Talamini, M.; Kavoussi, L.; Anderson, J. A telerobotic assistant for laparoscopic surgery. *IEEE Eng. Med. Biol. Mag.* **1995**, *14*, 279–288. <https://doi.org/10.1109/51.391776>.
25. Stoianovici, D.; Whitcomb, L.L.; Anderson, J.H.; Taylor, R.H.; Kavoussi, L.R. Modular surgical robotic system. In Proceedings of the International Conference on Medical Image Computing and Computer-Assisted Intervention, Cambridge, MA, USA, 11–13 October 1998. <https://doi.org/10.1007/bfb0056225>.
26. Lim, J.H.; Lee, W.J.; Park, D.W.; Yea, J.H.; Kim, S.H.; Kang, C.M. Revo-i robotic cholecystectomy study. *Surg. Endosc.* **2017**, *31*, 3391–3397. <https://doi.org/10.1007/s00464-016-5357-0>.
27. Dai, X.; Zhao, B.; Zhao, S.; He, Y.; Sun, Y.; Gao, P.; Zhang, J. Endoscope holder with tracking feature. In Proceedings of the 2016 IEEE International Conference on Information and Automation (ICIA), Ningbo, China, 1–3 August 2016; pp. 1303–1308. <https://doi.org/10.1109/ICInfA.2016.7831788>.
28. Guo, K.; Liu, Y.; Yang, H.; Zhao, S.; Zhang, Y.; Wang, Z.; Liu, B. Design and Fabrication of RCM structure used in Surgery Robot System. *J. Phys. Conf. Ser.* **2019**, *1345*, 052056. <https://doi.org/10.1088/1742-6596/1345/5/052056>.
29. Kastritsi, T.; Doulgeri, Z. A Controller to Impose a RCM for Hands-on Robotic-Assisted Minimally Invasive Surgery. *IEEE Trans. Med. Robot. Bionics* **2021**, *3*, 392–401. <https://doi.org/10.1109/TMRB.2021.3077319>.
30. Rao, P.P. New robotic platforms review. *World J. Urol.* **2018**, *36*, 537–541. <https://doi.org/10.1007/s00345-018-2213-y>.
31. Avatera Medical GmbH. Avatera System. Available online: <https://avatera.eu> (accessed on 11 May 2026).
32. Hagn, U.; Konietzschke, R.; Tobergte, A.; Nickl, M.; Jörg, S.; Kübler, B.; Passig, G.; Gröger, M.; Fröhlich, F.; Seibold, U.; et al. DLR MiroSurge system. *Int. J. Comput. Assist. Radiol. Surg.* **2010**, *5*, 183–193. <https://doi.org/10.1007/s11548-009-0372-4>.
33. Nasiri, E.; Sowrirajan, S.; Wang, L. Teleoperation in robot-assisted MIS with adaptive RCM via admittance control. *Int. J. Intell. Robot. Appl.* **2024**, *8*, 827–839. <https://doi.org/10.1007/s41315-024-00389-0>.
34. El Jjouaoui, H.; Cruz-Martinez, G.; Avila Vilchis, J.C.; Vilchis González, A.; Abdelaziz, S.; Poignet, P. Modeling of a Remote Center of Motion Spherical Parallel Tensegrity Mechanism for Percutaneous Interventions. In *Advances in Robot Kinematics*; Springer: Cham, Switzerland, 2022; pp. 327–336. https://doi.org/10.1007/978-3-031-08140-8_36.
35. Baumann, R.; Maeder, W.; Glauser, D.; Clavel, R. PantoScope mechanism. In Proceedings of the International Conference on Robotics and Automation, Albuquerque, NM, USA, 25 April 1997; pp. 718–723. <https://doi.org/10.1109/ROBOT.1997.620120>.
36. Hannaford, B.; Rosen, J.; Friedman, D.W.; King, H.; Roan, P.; Cheng, L.; Glozman, D.; Ma, J.; Kosari, S.N.; White, L. Raven-II platform. *IEEE Trans. Biomed. Eng.* **2013**, *60*, 954–959. <https://doi.org/10.1109/TBME.2012.2228858>.
37. Lum, M.J.H.; Rosen, J.; Sinanan, M.N.; Hannaford, B. Optimization of a Spherical Mechanism for a Minimally Invasive Surgical Robot. *IEEE Trans. Biomed. Eng.* **2006**, *53*, 1440–1445. <https://doi.org/10.1109/TBME.2006.875716>.
38. Essomba, T.; Nguyen Vu, L. Decoupled spherical RCM mechanism. *Mech. Mach. Theory* **2018**, *119*, 184–197. <https://doi.org/10.1016/j.mechmachtheory.2017.09.010>.
39. Vischer, P.; Clavel, R. Argos parallel wrist. *Int. J. Robot. Res.* **2000**, *19*, 5–11. <https://doi.org/10.1177/02783640022066707>.
40. Distalmotion SA. Dexter Surgical Robotic System—Technical Overview. Available online: <https://www.distalmotion.com> (accessed on 11 May 2026).
41. Distalmotion SA. Surgical Robotic System Comprising Spherical Wrist. U.S. Patent 12,402,961 B2, 2024.

42. Beira, R.; Santos-Carreras, L.; Rognini, G.; Clavel, R. Dionis RCM manipulator. *Appl. Bionics Biomech.* **2011**, *8*, 191–208. <https://doi.org/10.1155/2011/973097>.
43. Toshiba. TB67H420FTG Driver IC. Available online: <https://toshiba.semicon-storage.com> (accessed on 11 May 2026).
44. Allegro MicroSystems. ACS712 Sensor. Available online: <https://www.allegromicro.com> (accessed on 11 May 2026).
45. Wang, R.-J. Spherical Linkage Robotic Arm. U.S. Patent 9,440,362 B2, 2016.
46. Wang, R.-J. Surgical Robotic Arm. U.S. Patent 10,265,866 B2, 2019.
47. Riverfield Inc. Spherical Gear Mechanism. JP Patent 2022-106852 A.
48. Madhani, A.J.; Niemeyer, G.; Salisbury, J.K., Jr. The Black Falcon: A teleoperated surgical instrument for minimally invasive surgery. In Proceedings of the 1998 IEEE/RSJ International Conference on Intelligent Robots and Systems. Innovations in Theory, Practice and Applications (Cat. No.98CH36190), Victoria, BC, Canada, 17 October 1998; pp. 936–944. <https://doi.org/10.1109/IROS.1998.727320>.
49. Pisla, D.; Bulbucan, V.; Hedesiu, M.; Vaida, C.; Zima, I.; Mocan, R.; Tucan, P.; Dinu, C.; Pisla, D.; TEAM Project Group. A Vision-Guided Robotic System for Safe Dental Implant Surgery. *J. Clin. Med.* **2024**, *13*, 6326. <https://doi.org/10.3390/jcm13216326>.
50. Kuo, C.-H.; Dai, J.S. Fully-decoupled RCM manipulator. *ASME J. Med. Devices* **2012**, *6*, 021008. <https://doi.org/10.1115/1.4006541>.
51. Madhani, A.J.; Salisbury, J.K.; Niemeyer, G. Surgical Instrument Patent. U.S. Patent 5,797,900, 1998.
52. Intuitive Surgical. Surgical Tool Wrist Patent. U.S. Patent 12,390,294 B2, 2024.

Disclaimer/Publisher’s Note: The statements, opinions and data contained in all publications are solely those of the individual author(s) and contributor(s) and not of MDPI and/or the editor(s). MDPI and/or the editor(s) disclaim responsibility for any injury to people or property resulting from any ideas, methods, instructions or products referred to in the content.